

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122721\
 Data File : VX026073.D
 Acq On : 27 Dec 2021 19:36
 Operator : JC/MD
 Sample : M5174-06ME
 Misc : 5.66g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GBC17ME

Quant Time: Dec 28 01:33:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML122321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 23 21:00:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	193642	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	175270	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	89255	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.374	65	70238	51.150	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	102.300%
7) Chloroethane-d5	1.672	69	41246	37.585	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	75.180%
11) 1,1-Dichloroethene-d2	2.294	63	102937	37.828	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.660%
21) 2-Butanone-d5	4.489	46	105056	88.242	ug/L	0.03
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.240%
24) Chloroform-d	5.068	84	126545	46.050	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.100%
26) 1,2-Dichloroethane-d4	5.964	65	87596	47.080	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.160%
32) Benzene-d6	5.976	84	273171	54.148	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.300%
36) 1,2-Dichloropropane-d6	7.312	67	83419	53.195	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.380%
41) Toluene-d8	8.653	98	242343	51.449	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.900%
43) trans-1,3-Dichloroprop...	8.951	79	26043	38.171	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.340%
47) 2-Hexanone-d5	9.415	63	74574	92.143	ug/L	0.03
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.140%
56) 1,1,2,2-Tetrachloroeth...	11.201	84	101967	44.097	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.200%
66) 1,2-Dichlorobenzene-d4	12.323	152	86586	51.337	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122721\
Data File : VX026073.D
Acq On : 27 Dec 2021 19:36
Operator : JC/MD
Sample : M5174-06ME
Misc : 5.66g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GBC17ME

Quant Time: Dec 28 01:33:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM122321WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 23 21:00:32 2021
Response via : Initial Calibration

