

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122721\
 Data File : VX026076.D
 Acq On : 27 Dec 2021 21:31
 Operator : JC/MD
 Sample : VX1227WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK663

Quant Time: Dec 28 02:17:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML122321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 28 02:17:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	197611	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	178848	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	86219	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.374	65	74516	53.175	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	106.360%
7) Chloroethane-d5	1.673	69	57548	51.387	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	102.780%
11) 1,1-Dichloroethene-d2	2.313	63	104661	37.689	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.380%
21) 2-Butanone-d5	4.459	46	118741	97.734	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.730%
24) Chloroform-d	5.062	84	135042	48.155	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.300%
26) 1,2-Dichloroethane-d4	5.958	65	89552	47.165	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.320%
32) Benzene-d6	5.983	84	268990	52.253	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.500%
36) 1,2-Dichloropropane-d6	7.312	67	84646	52.897	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.800%
41) Toluene-d8	8.653	98	235812	49.061	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.120%
43) trans-1,3-Dichloroprop...	8.952	79	35593	51.125	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	102.240%
47) 2-Hexanone-d5	9.385	63	79204	95.905	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.910%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	111870	47.412	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.820%
66) 1,2-Dichlorobenzene-d4	12.323	152	83511	51.257	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.520%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122721\
Data File : VX026076.D
Acq On : 27 Dec 2021 21:31
Operator : JC/MD
Sample : VX1227WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK663

Quant Time: Dec 28 02:17:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM122321WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 28 02:17:03 2021
Response via : Initial Calibration

