

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122722\
 Data File : VX033561.D
 Acq On : 27 Dec 2022 13:16
 Operator : JC/MD
 Sample : VIBLK603
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK603

Quant Time: Dec 28 02:03:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML122722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 28 01:59:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	185047	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	164431	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	80299	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	39669	43.786	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	87.580%	
7) Chloroethane-d5	1.672	69	33266	38.985	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	77.960%	
11) 1,1-Dichloroethene-d2	2.306	63	54057	28.339	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	56.680%#	
21) 2-Butanone-d5	4.465	46	72538	107.313	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	107.310%	
24) Chloroform-d	5.056	84	113769	47.218	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	94.440%	
26) 1,2-Dichloroethane-d4	5.952	65	75905	50.096	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	100.200%	
32) Benzene-d6	5.970	84	218167	48.606	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.220%	
36) 1,2-Dichloropropane-d6	7.306	67	66790	50.411	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	100.820%	
41) Toluene-d8	8.647	98	208618	47.060	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	94.120%	
43) trans-1,3-Dichloroprop...	8.952	79	32328	49.792	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	99.580%	
47) 2-Hexanone-d5	9.384	63	55105	109.126	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	109.130%	
56) 1,1,2,2-Tetrachloroeth...	11.189	84	92700	50.566	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	101.140%	
66) 1,2-Dichlorobenzene-d4	12.323	152	81153	50.936	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	101.880%	
Target Compounds						Qvalue
72) 1,2,3-Trichlorobenzene	13.963	180	1037	0.678	ug/L	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122722\
Data File : VX033561.D
Acq On : 27 Dec 2022 13:16
Operator : JC/MD
Sample : VIBLK603
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK603

Quant Time: Dec 28 02:03:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM122722WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 28 01:59:02 2022
Response via : Initial Calibration

