

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122722\
 Data File : VX033567.D
 Acq On : 27 Dec 2022 15:34
 Operator : JC/MD
 Sample : VIBLK604
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK604

Quant Time: Dec 28 02:05:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML122722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 28 01:59:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	196850	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	173073	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	80852	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	50919	52.834	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	105.660%	
7) Chloroethane-d5	1.672	69	38938	42.896	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	85.800%	
11) 1,1-Dichloroethene-d2	2.306	63	62318	30.711	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	61.420%	
21) 2-Butanone-d5	4.465	46	73782	102.609	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	102.610%	
24) Chloroform-d	5.056	84	125340	48.901	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.800%	
26) 1,2-Dichloroethane-d4	5.958	65	79754	49.480	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	98.960%	
32) Benzene-d6	5.977	84	242787	51.390	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	102.780%	
36) 1,2-Dichloropropane-d6	7.306	67	72115	51.712	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	103.420%	
41) Toluene-d8	8.647	98	232302	49.786	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	99.580%	
43) trans-1,3-Dichloroprop...	8.952	79	34322	50.224	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	100.440%	
47) 2-Hexanone-d5	9.384	63	55535	104.486	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	104.490%	
56) 1,1,2,2-Tetrachloroeth...	11.189	84	94509	48.979	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	97.960%	
66) 1,2-Dichlorobenzene-d4	12.323	152	83835	52.259	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	104.520%	
Target Compounds						
14) Carbon disulfide	2.514	76	5551	1.757	ug/L #	92
62) 1,3,5-Trimethylbenzene	11.451	105	3224	0.702	ug/L	96
63) 1,2,4-Trimethylbenzene	11.756	105	3077	0.681	ug/L	96
64) 1,3-Dichlorobenzene	11.975	146	2748	1.117	ug/L	96
65) 1,4-Dichlorobenzene	12.043	146	3080	1.230	ug/L	98
67) 1,2-Dichlorobenzene	12.335	146	2224	0.904	ug/L #	78
70) 1,2,4-trichlorobenzene	13.591	180	3115	2.026	ug/L	94
72) 1,2,3-Trichlorobenzene	13.963	180	3073	1.994	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122722\
Data File : VX033567.D
Acq On : 27 Dec 2022 15:34
Operator : JC/MD
Sample : VIBLK604
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK604

Quant Time: Dec 28 02:05:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM122722WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 28 01:59:02 2022
Response via : Initial Calibration

