

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122724\
 Data File : VX044529.D
 Acq On : 27 Dec 2024 18:26
 Operator : JC/MD
 Sample : P5326-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VHBLK001

Quant Time: Dec 27 22:18:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120524WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 27 22:07:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.756	114	207194	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	177169	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	73665	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	77654	40.955	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.920%
7) Chloroethane-d5	1.642	69	58968	40.148	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	80.300%
11) 1,1-Dichloroethene-d2	2.294	65	31843	39.269	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.540%
21) 2-Butanone-d5	4.458	46	105702	91.054	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.050%
24) Chloroform-d	5.049	84	158773	44.103	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.200%
26) 1,2-Dichloroethane-d4	5.952	65	108305	42.541	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.080%
32) Benzene-d6	5.970	84	298822	48.813	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.620%
36) 1,2-Dichloropropane-d6	7.305	67	92231	48.521	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.040%
41) Toluene-d8	8.646	98	256009	45.229	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.460%
43) trans-1,3-Dichloroprop...	8.951	79	36193	39.621	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.240%
47) 2-Hexanone-d5	9.384	63	69021	93.126	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.130%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	121349	47.610	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.220%
66) 1,2-Dichlorobenzene-d4	12.316	152	87251	50.835	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.660%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122724\
Data File : VX044529.D
Acq On : 27 Dec 2024 18:26
Operator : JC/MD
Sample : P5326-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VHBLK001

Quant Time: Dec 27 22:18:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM120524WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 27 22:07:46 2024
Response via : Initial Calibration

