

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122818\
 Data File : VX006854.D
 Acq On : 28 Dec 2018 12:07
 Operator : JC/MD
 Sample : J6584-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RECLAIMER-PIT

Quant Time: Dec 29 05:50:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	706784	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1042683	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	970941	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	482300	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	350270	45.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.96%	
35) Dibromofluoromethane	5.51	113	309419	46.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.86%	
50) Toluene-d8	8.71	98	1229151	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
62) 4-Bromofluorobenzene	11.13	95	445091	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
Target Compounds						
16) Acetone	2.45	43	198494	56.718	ug/l	99
20) Methylene Chloride	2.86	84	3801366	525.288	ug/l	96
25) 2-Butanone	4.70	43	63610	12.922	ug/l	96
43) Isopropyl Acetate	6.46	43	24018	1.756	ug/l	98
51) 4-Methyl-2-Pentanone	8.65	43	170349	19.577	ug/l	93
52) Toluene	8.79	92	185425	10.934	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122818\
Data File : VX006854.D
Acq On : 28 Dec 2018 12:07
Operator : JC/MD
Sample : J6584-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RECLAIMER-PIT

Quant Time: Dec 29 05:50:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

