

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122818\
 Data File : VX006858.D
 Acq On : 28 Dec 2018 13:53
 Operator : JC/MD
 Sample : J6592-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 02 10:52:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	672040	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	996154	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	915425	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	442941	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	339675	46.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.80%	
35) Dibromofluoromethane	5.51	113	305382	48.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.96%	
50) Toluene-d8	8.71	98	1180059	49.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.38%	
62) 4-Bromofluorobenzene	11.13	95	421390	48.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.90%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122818\
Data File : VX006858.D
Acq On : 28 Dec 2018 13:53
Operator : JC/MD
Sample : J6592-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 02 10:52:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

