

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122818\
 Data File : VX006860.D
 Acq On : 28 Dec 2018 14:47
 Operator : JC/MD
 Sample : J6193-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-02-122718-D

Quant Time: Dec 29 05:48:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	666222	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1000791	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	931565	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	448961	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	340323	47.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.80%	
35) Dibromofluoromethane	5.51	113	302518	47.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.60%	
50) Toluene-d8	8.71	98	1194803	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
62) 4-Bromofluorobenzene	11.13	95	421937	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.03	59	6886	4.297	ug/l #	87
16) Acetone	2.45	43	76641	23.233	ug/l	95
20) Methylene Chloride	2.86	84	4087419	599.204	ug/l	95
25) 2-Butanone	4.70	43	6538	1.409	ug/l	97
29) Tetrahydrofuran	5.16	42	3859	1.263	ug/l #	85
43) Isopropyl Acetate	6.46	43	29109	2.218	ug/l	96
95) Naphthalene	13.83	128	74745	2.268	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122818\
Data File : VX006860.D
Acq On : 28 Dec 2018 14:47
Operator : JC/MD
Sample : J6193-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-02-122718-D

Quant Time: Dec 29 05:48:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

