

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122818\
 Data File : VX006866.D
 Acq On : 28 Dec 2018 17:27
 Operator : JC/MD
 Sample : J6486-02
 Misc : 6.29g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S72-0054

Quant Time: Dec 29 05:40:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	582563	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	862445	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	787647	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	416168	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	289466	46.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.20%	
35) Dibromofluoromethane	5.50	113	241990	44.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.74%	
50) Toluene-d8	8.71	98	1012000	49.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.44%	
62) 4-Bromofluorobenzene	11.14	95	373131	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
Target Compounds						
18) Methyl Acetate	2.80	43	38208	4.505	ug/l	98
20) Methylene Chloride	2.85	84	6870	1.152	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122818\
Data File : VX006866.D
Acq On : 28 Dec 2018 17:27
Operator : JC/MD
Sample : J6486-02
Misc : 6.29g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S72-0054

Quant Time: Dec 29 05:40:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

