

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122821\
 Data File : VX026104.D
 Acq On : 28 Dec 2021 17:41
 Operator : JC/MD
 Sample : VX1228MBL01
 Misc : 5.0g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK649

Quant Time: Dec 29 04:20:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML122821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 29 04:20:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	204078	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	177146	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	79729	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.374	65	91654	57.864	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	115.720%
7) Chloroethane-d5	1.672	69	69414	60.283	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	120.560%
11) 1,1-Dichloroethene-d2	2.313	63	130368	45.091	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.180%
21) 2-Butanone-d5	4.459	46	131021	119.708	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	119.710%
24) Chloroform-d	5.062	84	152636	56.742	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	113.480%
26) 1,2-Dichloroethane-d4	5.958	65	102641	58.864	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	117.720%
32) Benzene-d6	5.977	84	302284	59.188	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	118.380%
36) 1,2-Dichloropropane-d6	7.312	67	93594	59.239	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	118.480%
41) Toluene-d8	8.653	98	274840	59.007	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	118.020%
43) trans-1,3-Dichloroprop...	8.952	79	44789	58.073	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	116.140%
47) 2-Hexanone-d5	9.384	63	91607	124.322	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	124.320%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	122212	58.121	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	116.240%
66) 1,2-Dichlorobenzene-d4	12.323	152	90019	59.189	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	118.380%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122821\
Data File : VX026104.D
Acq On : 28 Dec 2021 17:41
Operator : JC/MD
Sample : VX1228MBL01
Misc : 5.0g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK649

Quant Time: Dec 29 04:20:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM122821WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 29 04:20:17 2021
Response via : Initial Calibration

