

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122821\
 Data File : VX026106.D
 Acq On : 28 Dec 2021 18:50
 Operator : JC/MD
 Sample : M5153-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VHBLK002

Quant Time: Dec 29 04:32:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML122821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 29 04:20:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	220762	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	191832	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	89324	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.374	65	88081	51.406	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 102.820%	
7) Chloroethane-d5	1.672	69	66887	53.699	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 107.400%	
11) 1,1-Dichloroethene-d2	2.313	63	123500	39.487	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 78.980%	
21) 2-Butanone-d5	4.452	46	129283	109.193	ug/L	-0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	= 109.190%	
24) Chloroform-d	5.056	84	145268	49.922	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 99.840%	
26) 1,2-Dichloroethane-d4	5.958	65	97723	51.808	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 103.620%	
32) Benzene-d6	5.976	84	291731	52.748	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 105.500%	
36) 1,2-Dichloropropane-d6	7.312	67	90362	52.815	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 105.620%	
41) Toluene-d8	8.653	98	261211	51.787	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 103.580%	
43) trans-1,3-Dichloroprop...	8.952	79	42435	50.809	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 101.620%	
47) 2-Hexanone-d5	9.384	63	86576	108.499	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 108.500%	
56) 1,1,2,2-Tetrachloroeth...	11.195	84	117583	51.638	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 103.280%	
66) 1,2-Dichlorobenzene-d4	12.323	152	89786	52.694	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 105.380%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122821\
Data File : VX026106.D
Acq On : 28 Dec 2021 18:50
Operator : JC/MD
Sample : M5153-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VHBLK002

Quant Time: Dec 29 04:32:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM122821WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 29 04:20:17 2021
Response via : Initial Calibration

