

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122821\
 Data File : VX026107.D
 Acq On : 28 Dec 2021 19:13
 Operator : JC/MD
 Sample : M5213-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VHBLK002

Quant Time: Dec 29 04:32:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML122821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 29 04:20:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	204565	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	178402	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	78822	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.374	65	86095	54.225	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	108.460%
7) Chloroethane-d5	1.673	69	66142	57.305	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	114.600%
11) 1,1-Dichloroethene-d2	2.313	63	122942	42.421	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.840%
21) 2-Butanone-d5	4.459	46	124740	113.698	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	113.700%
24) Chloroform-d	5.062	84	144345	53.533	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.060%
26) 1,2-Dichloroethane-d4	5.958	65	96021	54.936	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.880%
32) Benzene-d6	5.977	84	288607	56.112	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.220%
36) 1,2-Dichloropropane-d6	7.312	67	87529	55.010	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	110.020%
41) Toluene-d8	8.653	98	255183	54.401	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.800%
43) trans-1,3-Dichloroprop...	8.952	79	41642	53.613	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	107.220%
47) 2-Hexanone-d5	9.385	63	83437	112.437	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	112.440%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	113491	53.593	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	107.180%
66) 1,2-Dichlorobenzene-d4	12.323	152	85820	57.077	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	114.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122821\
Data File : VX026107.D
Acq On : 28 Dec 2021 19:13
Operator : JC/MD
Sample : M5213-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VHBLK002

Quant Time: Dec 29 04:32:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM122821WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 29 04:20:17 2021
Response via : Initial Calibration

