

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122823\
 Data File : VX039512.D
 Acq On : 28 Dec 2023 13:38
 Operator : JC/MD
 Sample : 06019-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Dec 28 22:49:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	78098	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	163417	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	157077	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	61374	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	68603	47.951	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.900%
35) Dibromofluoromethane	5.385	113	53342	47.249	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.500%
50) Toluene-d8	8.647	98	205612	49.057	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.120%
62) 4-Bromofluorobenzene	11.079	95	78342	45.302	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122823\
Data File : VX039512.D
Acq On : 28 Dec 2023 13:38
Operator : JC/MD
Sample : 06019-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Dec 28 22:49:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

