

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122823\
 Data File : VX039518.D
 Acq On : 28 Dec 2023 15:55
 Operator : JC/MD
 Sample : 06015-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MR-323-0-2

Quant Time: Dec 28 22:52:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	88951	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189270	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186782	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76729	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80134	49.177	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.360%
35) Dibromofluoromethane	5.385	113	61980	47.401	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.800%
50) Toluene-d8	8.647	98	237413	48.907	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.820%
62) 4-Bromofluorobenzene	11.079	95	94420	47.142	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.280%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	42604	23.818	ug/l	96
20) Methylene Chloride	2.788	84	7130	5.249	ug/l	93
37) Ethyl Acetate	4.733	43	3076	1.207	ug/l #	83
43) Isopropyl Acetate	6.342	43	111307	26.851	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122823\
Data File : VX039518.D
Acq On : 28 Dec 2023 15:55
Operator : JC/MD
Sample : 06015-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MR-323-0-2

Quant Time: Dec 28 22:52:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

