

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122823\
 Data File : VX039519.D
 Acq On : 28 Dec 2023 16:18
 Operator : JC/MD
 Sample : 06015-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MR-323-6-8

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/29/2023
 Supervised By : Semsettin Yesilyurt 01/01/2024

Quant Time: Dec 28 22:52:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	92516	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	194817	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	194292	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80736	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	82911	48.920	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.840%		
35) Dibromofluoromethane	5.385	113	62912	46.744	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.480%		
50) Toluene-d8	8.647	98	251583	50.350	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	100.700%		
62) 4-Bromofluorobenzene	11.079	95	98528	47.792	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.580%		
Target Compounds					Qvalue	
16) Acetone	2.386	43	42823	22.157	ug/l	97
20) Methylene Chloride	2.788	84	7038	4.981	ug/l #	91
37) Ethyl Acetate	4.733	43	2987m	1.138	ug/l	
43) Isopropyl Acetate	6.342	43	112538	26.375	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122823\
Data File : VX039519.D
Acq On : 28 Dec 2023 16:18
Operator : JC/MD
Sample : 06015-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MR-323-6-8

Manual Integrations
APPROVED

Reviewed By : John Carlone 12/29/2023
Supervised By : Semsettin Yesilyurt 01/01/2024

Quant Time: Dec 28 22:52:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

