

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122823\
 Data File : VX039520.D
 Acq On : 28 Dec 2023 16:40
 Operator : JC/MD
 Sample : 06016-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MR-322-0-2

Quant Time: Dec 28 22:53:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/29/2023
 Supervised By : Semsettin Yesilyurt 01/01/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	88925	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	185674	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	187155	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77088	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	78965	48.473	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.940%		
35) Dibromofluoromethane	5.385	113	60991	47.548	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.100%		
50) Toluene-d8	8.647	98	238332	50.047	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	100.100%		
62) 4-Bromofluorobenzene	11.079	95	93371	47.521	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.040%		
Target Compounds						
11) Tert butyl alcohol	2.947	59	5171	12.641	ug/l #	81
16) Acetone	2.386	43	43874	25.307	ug/l	97
20) Methylene Chloride	2.788	84	6972	5.134	ug/l	91
37) Ethyl Acetate	4.727	43	3252m	1.300	ug/l	
43) Isopropyl Acetate	6.342	43	110476	27.166	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122823\
Data File : VX039520.D
Acq On : 28 Dec 2023 16:40
Operator : JC/MD
Sample : 06016-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MR-322-0-2

Quant Time: Dec 28 22:53:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/29/2023
Supervised By :Semsettin Yesilyurt 01/01/2024

