

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122823\
 Data File : VX039523.D
 Acq On : 28 Dec 2023 17:49
 Operator : JC/MD
 Sample : 05978-01
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 N41547

Manual Integrations
 APPROVED

Reviewed By :John Carlone 12/29/2023
 Supervised By :Semsettin Yesilyurt 01/01/2024

Quant Time: Dec 28 18:16:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	81736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	169550	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	162643	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63678	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71158	47.523	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	95.040%		
35) Dibromofluoromethane	5.391	113	52636	44.937	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	89.880%		
50) Toluene-d8	8.647	98	213517	49.100	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.200%		
62) 4-Bromofluorobenzene	11.085	95	79725	44.434	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	88.860%		
Target Compounds						
18) Methyl Acetate	2.733	43	3406m	1.052	ug/l	
84) 1,2,4-Trimethylbenzene	11.756	105	1259	0.293	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122823\
Data File : VX039523.D
Acq On : 28 Dec 2023 17:49
Operator : JC/MD
Sample : 05978-01
Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
N41547

Quant Time: Dec 28 18:16:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 12/29/2023
Supervised By : Semsettin Yesilyurt 01/01/2024

