

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122920\
 Data File : VX020426.D
 Acq On : 29 Dec 2020 11:59
 Operator : JC/MD
 Sample : VX1229WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1229WBSD01

Quant Time: Dec 30 02:35:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	226929	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.824	114	356772	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	336680	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	166300	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	140098	46.06	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.12%
35) Dibromofluoromethane	5.458	113	115280	50.51	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.02%
50) Toluene-d8	8.696	98	442201	50.34	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.68%
62) 4-Bromofluorobenzene	11.122	95	170089	50.59	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.18%

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.185	85	39971	16.22	ug/l 99
3) Chloromethane	1.313	50	45257	15.84	ug/l 97
4) Vinyl Chloride	1.392	62	49575	15.92	ug/l 99
5) Bromomethane	1.624	94	18430	18.48	ug/l 98
6) Chloroethane	1.703	64	33261	17.16	ug/l 96
7) Trichlorofluoromethane	1.910	101	69274	17.30	ug/l 99
8) Diethyl Ether	2.166	74	32365	18.58	ug/l 99
9) 1,1,2-Trichlorotrifluo...	2.361	101	41626	18.37	ug/l 99
10) Methyl Iodide	2.489	142	40630	15.72	ug/l 99
11) Tert butyl alcohol	3.014	59	70332	103.25	ug/l 100
12) 1,1-Dichloroethene	2.355	96	40170	17.33	ug/l 98
13) Acrolein	2.270	56	40142	94.31	ug/l 99
14) Allyl chloride	2.703	41	77775	18.90	ug/l 91
15) Acrylonitrile	3.111	53	165517	105.62	ug/l 99
16) Acetone	2.416	43	119390	86.07	ug/l 98
17) Carbon Disulfide	2.544	76	99002	14.46	ug/l 100
18) Methyl Acetate	2.745	43	64719	20.92	ug/l 100
19) Methyl tert-butyl Ether	3.166	73	172763	19.91	ug/l 99
20) Methylene Chloride	2.831	84	56063	17.93	ug/l 99
21) trans-1,2-Dichloroethene	3.136	96	49116	17.46	ug/l 97
22) Diisopropyl ether	3.825	45	170195	20.27	ug/l 90
23) Vinyl Acetate	3.782	43	713770	99.73	ug/l 100
24) 1,1-Dichloroethane	3.666	63	95401	19.29	ug/l 99
25) 2-Butanone	4.629	43	206410	102.24	ug/l 100
26) 2,2-Dichloropropane	4.544	77	74522	18.26	ug/l 100
27) cis-1,2-Dichloroethene	4.556	96	59382	18.70	ug/l 99
28) Bromochloromethane	4.977	49	38242	17.36	ug/l 97
29) Tetrahydrofuran	5.086	42	132038	103.96	ug/l 98
30) Chloroform	5.172	83	97151	19.29	ug/l 99
31) Cyclohexane	5.544	56	75493	18.00	ug/l 96
32) 1,1,1-Trichloroethane	5.458	97	79616	18.96	ug/l 99
36) 1,1-Dichloropropene	5.763	75	66351	18.82	ug/l 99
37) Ethyl Acetate	4.788	43	79461	21.69	ug/l 100
38) Carbon Tetrachloride	5.751	117	66889	19.95	ug/l 97
39) Methylcyclohexane	7.434	83	77129	18.95	ug/l 97
40) Benzene	6.105	78	217233	19.79	ug/l 100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122920\
 Data File : VX020426.D
 Acq On : 29 Dec 2020 11:59
 Operator : JC/MD
 Sample : VX1229WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1229WBSD01

Quant Time: Dec 30 02:35:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.995	41	42875	21.96	ug/l	98
42) 1,2-Dichloroethane	6.153	62	73929	19.64	ug/l	98
43) Isopropyl Acetate	6.409	43	123434	21.34	ug/l	99
44) Trichloroethene	7.184	130	55294	19.49	ug/l	100
45) 1,2-Dichloropropane	7.482	63	57841	20.85	ug/l	98
46) Dibromomethane	7.635	93	37713	19.88	ug/l	99
47) Bromodichloromethane	7.873	83	74756	20.67	ug/l	98
48) Methyl methacrylate	7.745	41	59921	21.14	ug/l	95
49) 1,4-Dioxane	7.714	88	29354	442.36	ug/l	97
51) 4-Methyl-2-Pentanone	8.616	43	399954	111.68	ug/l	100
52) Toluene	8.763	92	137502	20.53	ug/l	97
53) t-1,3-Dichloropropene	9.019	75	81771	20.55	ug/l	99
54) cis-1,3-Dichloropropene	8.415	75	90797	20.78	ug/l	92
55) 1,1,2-Trichloroethane	9.195	97	59293	21.46	ug/l	98
56) Ethyl methacrylate	9.159	69	87301	21.59	ug/l	97
57) 1,3-Dichloropropane	9.348	76	98996	21.05	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.287	63	194595	90.99	ug/l	99
59) 2-Hexanone	9.470	43	301724	110.56	ug/l	99
60) Dibromochloromethane	9.561	129	59230	21.61	ug/l	100
61) 1,2-Dibromoethane	9.653	107	59807	20.73	ug/l	99
64) Tetrachloroethene	9.317	164	53164	20.96	ug/l	95
65) Chlorobenzene	10.116	112	144525	20.01	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.201	131	54229	20.92	ug/l	99
67) Ethyl Benzene	10.232	91	255181	20.21	ug/l	99
68) m/p-Xylenes	10.342	106	194317	41.22	ug/l	99
69) o-Xylene	10.683	106	92097	20.30	ug/l	100
70) Styrene	10.695	104	161759	21.09	ug/l	100
71) Bromoform	10.835	173	43005	21.42	ug/l #	99
73) Isopropylbenzene	11.000	105	254775	21.23	ug/l	100
74) N-amyl acetate	10.878	43	98895	20.44	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.250	83	92381	21.19	ug/l	99
76) 1,2,3-Trichloropropane	11.281	75	105778	25.82	ug/l	87
77) Bromobenzene	11.238	156	64562	20.35	ug/l	100
78) n-propylbenzene	11.341	91	289945	20.96	ug/l	99
79) 2-Chlorotoluene	11.402	91	177258	20.55	ug/l	100
80) 1,3,5-Trimethylbenzene	11.488	105	216739	21.30	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.055	75	28578	20.70	ug/l	98
82) 4-Chlorotoluene	11.494	91	207979	20.64	ug/l	100
83) tert-Butylbenzene	11.756	119	202606	20.75	ug/l	97
84) 1,2,4-Trimethylbenzene	11.793	105	218603	21.17	ug/l	99
85) sec-Butylbenzene	11.927	105	240956	21.39	ug/l	100
86) p-Isopropyltoluene	12.049	119	222795	21.44	ug/l	99
87) 1,3-Dichlorobenzene	12.006	146	116510	20.54	ug/l	99
88) 1,4-Dichlorobenzene	12.079	146	116499	19.69	ug/l	97
89) n-Butylbenzene	12.372	91	190870	20.97	ug/l	99
90) Hexachloroethane	12.579	117	36777	20.52	ug/l	96
91) 1,2-Dichlorobenzene	12.372	146	116457	20.63	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.981	75	18963	20.73	ug/l	100
93) 1,2,4-Trichlorobenzene	13.628	180	78040	21.50	ug/l	99
94) Hexachlorobutadiene	13.762	225	36185	25.14	ug/l	99
95) Naphthalene	13.817	128	249477	21.71	ug/l	99
96) 1,2,3-Trichlorobenzene	14.000	180	76796	21.39	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122920\
Data File : VX020426.D
Acq On : 29 Dec 2020 11:59
Operator : JC/MD
Sample : VX1229WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01

Quant Time: Dec 30 02:35:02 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.623 min Scan# 744
Delta R.T. 0.006 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion:168 Resp: 226929
Ion Ratio Lower Upper
168 100
99 55.7 45.8 68.6



#2
Dichlorodifluoromethane
Concen: 16.22 ug/l
RT: 1.185 min Scan# 16
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 85 Resp: 39971
Ion Ratio Lower Upper
85 100
87 32.5 16.4 49.2









#3
Chloromethane
Concen: 15.84 ug/l
RT: 1.313 min Scan# 37
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 50 Resp: 45257
Ion Ratio Lower Upper
50 100
52 34.4 26.4 39.6



#4
Vinyl Chloride
Concen: 15.92 ug/l
RT: 1.392 min Scan# 50
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 62 Resp: 49575
Ion Ratio Lower Upper
62 100
64 32.7 25.7 38.5









#5
Bromomethane
Concen: 18.48 ug/l
RT: 1.624 min Scan# 88
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 94 Resp: 18430
Ion Ratio Lower Upper
94 100
96 97.0 76.4 114.6



#6
Chloroethane
Concen: 17.16 ug/l
RT: 1.703 min Scan# 101
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 64 Resp: 33261
Ion Ratio Lower Upper
64 100
66 33.7 25.4 38.0









#7
 Trichlorofluoromethane
 Concen: 17.30 ug/l
 RT: 1.910 min Scan# 135
 Delta R.T. 0.000 min
 Lab File: VX020426.D
 Acq: 29 Dec 2020 11:59




Instrument :
 MSVOA_X
 ClientSampleId :
 VX1229WBSD01


Tgt Ion:101 Resp: 69274
 Ion Ratio Lower Upper
 101 100
 103 67.2 53.2 79.8



#8
 Diethyl Ether
 Concen: 18.58 ug/l
 RT: 2.166 min Scan# 177
 Delta R.T. 0.000 min
 Lab File: VX020426.D
 Acq: 29 Dec 2020 11:59

Tgt Ion: 74 Resp: 32365
 Ion Ratio Lower Upper
 74 100
 45 89.7 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.37 ug/l
RT: 2.361 min Scan# 209
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion:101 Resp: 41626
Ion Ratio Lower Upper
101 100
85 41.8 34.9 52.3
151 76.6 61.4 92.2



#10
Methyl Iodide
Concen: 15.72 ug/l
RT: 2.489 min Scan# 230
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion:142 Resp: 40630
Ion Ratio Lower Upper
142 100
127 41.9 33.3 49.9
141 14.4 11.4 17.0









#11
Tert butyl alcohol
Concen: 103.25 ug/l
RT: 3.014 min Scan# 316
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 59 Resp: 70332
Ion Ratio Lower Upper
59 100
57 9.8 7.8 11.8



#12
1,1-Dichloroethene
Concen: 17.33 ug/l
RT: 2.355 min Scan# 208
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 96 Resp: 40170
Ion Ratio Lower Upper
96 100
61 154.2 124.6 187.0
98 60.6 50.6 76.0









#13
Acrolein
Concen: 94.31 ug/l
RT: 2.270 min Scan# 194
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 56 Resp: 40142
Ion Ratio Lower Upper
56 100
55 71.7 56.4 84.6



#14
Allyl chloride
Concen: 18.90 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 41 Resp: 77775
Ion Ratio Lower Upper
41 100
39 56.0 51.6 77.4
76 35.2 30.5 45.7









#15
Acrylonitrile
Concen: 105.62 ug/l
RT: 3.111 min Scan# 332
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 53 Resp: 165517
Ion Ratio Lower Upper
53 100
52 82.1 65.3 97.9
51 34.7 28.2 42.2



#16
Acetone
Concen: 86.07 ug/l
RT: 2.416 min Scan# 218
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 43 Resp: 119390
Ion Ratio Lower Upper
43 100
58 35.9 27.8 41.8









#17
Carbon Disulfide
Concen: 14.46 ug/l
RT: 2.544 min Scan# 239
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 76 Resp: 99002
Ion Ratio Lower Upper
76 100
78 9.0 7.2 10.8



#18
Methyl Acetate
Concen: 20.92 ug/l
RT: 2.745 min Scan# 272
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 43 Resp: 64719
Ion Ratio Lower Upper
43 100
74 27.2 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 19.91 ug/l
RT: 3.166 min Scan# 341
Delta R.T. 0.006 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 73 Resp: 172763
Ion Ratio Lower Upper
73 100
57 21.1 17.1 25.7



#20
Methylene Chloride
Concen: 17.93 ug/l
RT: 2.831 min Scan# 286
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59


Tgt Ion: 84 Resp: 56063
Ion Ratio Lower Upper
84 100
49 109.5 88.6 132.8
51 33.9 27.0 40.6
86 60.7 50.3 75.5








#21
trans-1,2-Dichloroethene
Concen: 17.46 ug/l
RT: 3.136 min Scan# 336
Delta R.T. 0.001 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59


Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01





Tgt Ion: 96 Resp: 49116
Ion Ratio Lower Upper
96 100
61 138.4 107.7 161.5
98 60.5 50.4 75.6




#22
Diisopropyl ether
Concen: 20.27 ug/l
RT: 3.825 min Scan# 449
Delta R.T. 0.007 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59



Tgt Ion: 45 Resp: 170195
Ion Ratio Lower Upper
45 100
43 61.1 58.0 87.0
87 28.9 25.4 38.0
59 12.9 10.2 15.2









#23
 Vinyl Acetate
 Concen: 99.73 ug/l
 RT: 3.782 min Scan# 442
 Delta R.T. 0.006 min
 Lab File: VX020426.D
 Acq: 29 Dec 2020 11:59




Instrument :
 MSVOA_X
 ClientSampleId :
 VX1229WBSD01


Tgt Ion: 43 Resp: 713770
 Ion Ratio Lower Upper
 43 100
 86 12.4 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 19.29 ug/l
 RT: 3.666 min Scan# 423
 Delta R.T. 0.000 min
 Lab File: VX020426.D
 Acq: 29 Dec 2020 11:59

Tgt Ion: 63 Resp: 95401
 Ion Ratio Lower Upper
 63 100
 98 7.8 3.8 11.2
 100 4.5 2.5 7.5









#25
2-Butanone
Concen: 102.24 ug/l
RT: 4.629 min Scan# 581
Delta R.T. 0.006 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 43 Resp: 206410
Ion Ratio Lower Upper
43 100
72 29.4 23.7 35.5



#26
2,2-Dichloropropane
Concen: 18.26 ug/l
RT: 4.544 min Scan# 567
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 77 Resp: 74522
Ion Ratio Lower Upper
77 100
97 23.3 11.7 35.1





#27

cis-1,2-Dichloroethene

Concen: 18.70 ug/l

RT: 4.556 min Scan# 569

Delta R.T. 0.000 min

Lab File: VX020426.D

Acq: 29 Dec 2020 11:59

Instrument :

MSVOA_X

ClientSampleId :

VX1229WBSD01





Tgt Ion: 96 Resp: 59382

Ion Ratio Lower Upper

96 100

61 141.3 0.0 285.6

98 64.2 0.0 129.6



#28

Bromochloromethane

Concen: 17.36 ug/l

RT: 4.977 min Scan# 638

Delta R.T. 0.007 min

Lab File: VX020426.D

Acq: 29 Dec 2020 11:59




Tgt Ion: 49 Resp: 38242


Ion Ratio Lower Upper

49 100

129 1.9 0.0 4.0

130 82.5 68.2 102.2





#29

Tetrahydrofuran

Concen: 103.96 ug/l

RT: 5.086 min Scan# 656





Delta R.T. 0.006 min

Lab File: VX020426.D

Acq: 29 Dec 2020 11:59

Tgt Ion: 42 Resp: 132038

Ion	Ratio	Lower	Upper
42	100		
72	50.8	41.7	62.5
71	47.1	38.4	57.6



#30

Chloroform

Concen: 19.29 ug/l

RT: 5.172 min Scan# 670




Delta R.T. -0.000 min


Lab File: VX020426.D

Acq: 29 Dec 2020 11:59

Tgt Ion: 83 Resp: 97151


Ion	Ratio	Lower	Upper
83	100		
85	65.7	52.0	78.0








#31
Cyclohexane
Concen: 18.00 ug/l
RT: 5.544 min Scan# 731
Delta R.T. 0.007 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59


Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01





Tgt Ion: 56 Resp: 75493
Ion Ratio Lower Upper
56 100
69 31.9 25.3 37.9
84 84.5 71.4 107.0




#32
1,1,1-Trichloroethane
Concen: 18.96 ug/l
RT: 5.458 min Scan# 717
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59



Tgt Ion: 97 Resp: 79616
Ion Ratio Lower Upper
97 100
99 63.8 51.5 77.3
61 41.8 34.2 51.4









#33
1,2-Dichloroethane-d4
Concen: 46.06 ug/l
RT: 6.025 min Scan# 810
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 65 Resp: 140098
Ion Ratio Lower Upper
65 100
67 54.1 0.0 105.4



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 6.824 min Scan# 941
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 114 Resp: 356772
Ion Ratio Lower Upper
114 100
63 20.6 0.0 40.8
88 16.7 0.0 32.8









#35
Dibromofluoromethane
Concen: 50.51 ug/l
RT: 5.458 min Scan# 717
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion:113 Resp: 115280
Ion Ratio Lower Upper
113 100
111 102.6 81.9 122.9
192 21.5 16.7 25.1



#36
1,1-Dichloropropene
Concen: 18.82 ug/l
RT: 5.763 min Scan# 767
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 75 Resp: 66351
Ion Ratio Lower Upper
75 100
110 35.8 18.1 54.4
77 31.4 25.3 37.9





#37

Ethyl Acetate

Concen: 21.69 ug/l

RT: 4.788 min Scan# 607





Delta R.T. -0.000 min

Lab File: VX020426.D

Acq: 29 Dec 2020 11:59

Tgt Ion:	43	Resp:	79461
Ion	Ratio	Lower	Upper
43	100		
61	15.9	12.8	19.2
70	13.1	10.5	15.7

Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01



#38

Carbon Tetrachloride

Concen: 19.95 ug/l




RT: 5.751 min Scan# 765


Delta R.T. 0.006 min

Lab File: VX020426.D

Acq: 29 Dec 2020 11:59

Tgt Ion:	117	Resp:	66889
Ion	Ratio	Lower	Upper
117	100		
119	98.4	76.2	114.2
121	30.5	24.8	37.2





#39

Methylcyclohexane

Concen: 18.95 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. 0.001 min

Lab File: VX020426.D

Acq: 29 Dec 2020 11:59





Tgt Ion: 83 Resp: 77129

Ion Ratio Lower Upper

83 100

55 76.1 58.9 88.3

98 48.9 38.1 57.1



#40

Benzene

Concen: 19.79 ug/l

RT: 6.105 min Scan# 823

Delta R.T. 0.001 min

Lab File: VX020426.D




Acq: 29 Dec 2020 11:59


Tgt Ion: 78 Resp: 217233

Ion Ratio Lower Upper

78 100

77 23.4 18.9 28.3









#41
Methacrylonitrile
Concen: 21.96 ug/l
RT: 4.995 min Scan# 641
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion:	41	Resp:	42875
Ion	Ratio	Lower	Upper
41	100		
39	50.8	42.5	63.7
67	80.6	65.0	97.6
52	30.0	25.3	37.9



#42
1,2-Dichloroethane
Concen: 19.64 ug/l
RT: 6.153 min Scan# 831
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion:	62	Resp:	73929
Ion	Ratio	Lower	Upper
62	100		
98	10.7	0.0	20.2









#43
Isopropyl Acetate
Concen: 21.34 ug/l
RT: 6.409 min Scan# 873
Delta R.T. 0.006 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 43 Resp: 123434
Ion Ratio Lower Upper
43 100
61 23.0 18.2 27.4
87 15.1 12.5 18.7



#44
Trichloroethene
Concen: 19.49 ug/l
RT: 7.184 min Scan# 1000
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion:130 Resp: 55294
Ion Ratio Lower Upper
130 100
95 98.7 0.0 198.0









#45
1,2-Dichloropropane
Concen: 20.85 ug/l
RT: 7.482 min Scan# 1049
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 63 Resp: 57841
Ion Ratio Lower Upper
63 100
65 30.8 25.7 38.5



#46
Dibromomethane
Concen: 19.88 ug/l
RT: 7.635 min Scan# 1074
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 93 Resp: 37713
Ion Ratio Lower Upper
93 100
95 82.1 66.8 100.2
174 104.9 83.1 124.7









#47
Bromodichloromethane
Concen: 20.67 ug/l
RT: 7.873 min Scan# 1113
Delta R.T. 0.001 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 83	Resp: 74756
Ion Ratio	Lower Upper
83 100	
85 65.9	51.4 77.0
127 8.9	6.8 10.2



#48
Methyl methacrylate
Concen: 21.14 ug/l
RT: 7.745 min Scan# 1092
Delta R.T. 0.007 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 41	Resp: 59921
Ion Ratio	Lower Upper
41 100	
69 96.6	76.0 114.0
39 48.9	44.6 67.0









#49
1,4-Dioxane
Concen: 442.36 ug/l
RT: 7.714 min Scan# 1087
Delta R.T. 0.006 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampled :
VX1229WBSD01


Tgt Ion: 88 Resp: 29354
Ion Ratio Lower Upper
88 100
43 28.8 23.2 34.8
58 72.1 55.0 82.4



#50
Toluene-d8
Concen: 50.34 ug/l
RT: 8.696 min Scan# 1248
Delta R.T. 0.007 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 98 Resp: 442201
Ion Ratio Lower Upper
98 100
100 66.0 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 111.68 ug/l
RT: 8.616 min Scan# 1235
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 43 Resp: 399954
Ion Ratio Lower Upper
43 100
58 44.2 35.4 53.0



#52
Toluene
Concen: 20.53 ug/l
RT: 8.763 min Scan# 1259
Delta R.T. 0.001 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 92 Resp: 137502
Ion Ratio Lower Upper
92 100
91 167.8 137.1 205.7









#53
 t-1,3-Dichloropropene
 Concen: 20.55 ug/l
 RT: 9.019 min Scan# 1301
 Delta R.T. -0.000 min
 Lab File: VX020426.D
 Acq: 29 Dec 2020 11:59




Instrument :
 MSVOA_X
 ClientSampleId :
 VX1229WBSD01


Tgt Ion: 75 Resp: 81771
 Ion Ratio Lower Upper
 75 100
 77 31.4 24.9 37.3



#54
 cis-1,3-Dichloropropene
 Concen: 20.78 ug/l
 RT: 8.415 min Scan# 1202
 Delta R.T. 0.006 min
 Lab File: VX020426.D
 Acq: 29 Dec 2020 11:59

Tgt Ion: 75 Resp: 90797
 Ion Ratio Lower Upper
 75 100
 77 31.3 25.7 38.5
 39 36.0 35.1 52.7









#55
1,1,2-Trichloroethane
Concen: 21.46 ug/l
RT: 9.195 min Scan# 1330
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion:	97	Resp:	59293
Ion	Ratio	Lower	Upper
97	100		
83	84.5	69.3	103.9
85	58.2	44.3	66.5
99	63.1	50.3	75.5



#56
Ethyl methacrylate
Concen: 21.59 ug/l
RT: 9.159 min Scan# 1324
Delta R.T. 0.006 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion:	69	Resp:	87301
Ion	Ratio	Lower	Upper
69	100		
41	58.3	46.8	70.2
39	27.0	24.7	37.1









#57
1,3-Dichloropropane
Concen: 21.05 ug/l
RT: 9.348 min Scan# 1355
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 76 Resp: 98996
Ion Ratio Lower Upper
76 100
78 32.2 26.2 39.2



#58
2-Chloroethyl Vinyl ether
Concen: 90.99 ug/l
RT: 8.287 min Scan# 1181
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 63 Resp: 194595
Ion Ratio Lower Upper
63 100
106 29.1 23.5 35.3









#59
2-Hexanone
Concen: 110.56 ug/l
RT: 9.470 min Scan# 1375
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 43 Resp: 301724
Ion Ratio Lower Upper
43 100
58 61.1 30.9 92.8



#60
Dibromochloromethane
Concen: 21.61 ug/l
RT: 9.561 min Scan# 1390
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 129 Resp: 59230
Ion Ratio Lower Upper
129 100
127 76.6 38.4 115.2









#61
1,2-Dibromoethane
Concen: 20.73 ug/l
RT: 9.653 min Scan# 1405
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion:107 Resp: 59807
Ion Ratio Lower Upper
107 100
109 95.9 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 50.59 ug/l
RT: 11.122 min Scan# 1646
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 95 Resp: 170089
Ion Ratio Lower Upper
95 100
174 75.9 0.0 154.6
176 75.8 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.092 min Scan# 1477
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion:117 Resp: 336680
Ion Ratio Lower Upper
117 100
82 55.8 47.4 71.2
119 31.6 26.6 39.8



#64
Tetrachloroethene
Concen: 20.96 ug/l
RT: 9.317 min Scan# 1350
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion:164 Resp: 53164
Ion Ratio Lower Upper
164 100
166 132.7 103.4 155.2
129 97.2 70.8 106.2
131 94.4 72.5 108.7









#65
Chlorobenzene
Concen: 20.01 ug/l
RT: 10.116 min Scan# 1481
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion:112 Resp: 144525
Ion Ratio Lower Upper
112 100
114 32.8 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 20.92 ug/l
RT: 10.201 min Scan# 1495
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion:131 Resp: 54229
Ion Ratio Lower Upper
131 100
133 97.0 47.7 143.1
119 65.2 32.1 96.5









#67
Ethyl Benzene
Concen: 20.21 ug/l
RT: 10.232 min Scan# 1500
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 91 Resp: 255181
Ion Ratio Lower Upper
91 100
106 30.5 24.6 37.0



#68
m/p-Xylenes
Concen: 41.22 ug/l
RT: 10.342 min Scan# 1518
Delta R.T. 0.001 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 106 Resp: 194317
Ion Ratio Lower Upper
106 100
91 205.2 162.8 244.2









#69
o-Xylene
Concen: 20.30 ug/l
RT: 10.683 min Scan# 1574
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion:106 Resp: 92097
Ion Ratio Lower Upper
106 100
91 216.5 108.1 324.2



#70
Styrene
Concen: 21.09 ug/l
RT: 10.695 min Scan# 1576
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion:104 Resp: 161759
Ion Ratio Lower Upper
104 100
78 49.7 40.0 60.0
103 53.8 43.4 65.0









#71
Bromoform
Concen: 21.42 ug/l
RT: 10.835 min Scan# 1599
Delta R.T. -0.006 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion:173 Resp: 43005
Ion Ratio Lower Upper
173 100
175 48.9 23.9 71.8
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.061 min Scan# 1800
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion:152 Resp: 166300
Ion Ratio Lower Upper
152 100
115 70.8 41.1 123.3
150 168.3 0.0 349.0









#73
Isopropylbenzene
Concen: 21.23 ug/l
RT: 11.000 min Scan# 1626
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion:105 Resp: 254775
Ion Ratio Lower Upper
105 100
120 26.2 13.2 39.6



#74
N-ethyl acetate
Concen: 20.44 ug/l
RT: 10.878 min Scan# 1606
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 43 Resp: 98895
Ion Ratio Lower Upper
43 100
70 50.0 39.0 58.6
55 28.4 21.5 32.3
61 28.3 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 21.19 ug/l
RT: 11.250 min Scan# 1667
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 83 Resp: 92381
Ion Ratio Lower Upper
83 100
131 9.8 5.1 15.2
85 63.0 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 25.82 ug/l
RT: 11.281 min Scan# 1672
Delta R.T. 0.001 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 75 Resp: 105778
Ion Ratio Lower Upper
75 100
77 30.4 19.1 57.5





#77

Bromobenzene

Concen: 20.35 ug/l

RT: 11.238 min Scan# 1665





Delta R.T. -0.000 min

Lab File: VX020426.D

Acq: 29 Dec 2020 11:59

Tgt Ion:	156	Resp:	64562
Ion Ratio	Lower	Upper	
156	100		
77	166.5	82.8	248.6
158	98.4	49.1	147.4

Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01



#78

n-propylbenzene

Concen: 20.96 ug/l




RT: 11.341 min Scan# 1682


Delta R.T. 0.000 min

Lab File: VX020426.D

Acq: 29 Dec 2020 11:59

Tgt Ion:	91	Resp:	289945
Ion Ratio	Lower	Upper	
91	100		
120	22.7	11.5	34.5









#79
2-Chlorotoluene
Concen: 20.55 ug/l
RT: 11.402 min Scan# 1692
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 91 Resp: 177258
Ion Ratio Lower Upper
91 100
126 33.1 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 21.30 ug/l
RT: 11.488 min Scan# 1706
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 105 Resp: 216739
Ion Ratio Lower Upper
105 100
120 48.0 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 20.70 ug/l
RT: 11.055 min Scan# 1635
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 75 Resp: 28578
Ion Ratio Lower Upper
75 100
53 94.0 74.5 111.7
89 45.4 37.9 56.9



#82
4-Chlorotoluene
Concen: 20.64 ug/l
RT: 11.494 min Scan# 1707
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 91 Resp: 207979
Ion Ratio Lower Upper
91 100
126 29.5 14.9 44.5









#83
tert-Butylbenzene
Concen: 20.75 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion:119 Resp: 202606
Ion Ratio Lower Upper
119 100
91 58.2 27.9 83.7
134 23.5 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 21.17 ug/l
RT: 11.793 min Scan# 1756
Delta R.T. 0.001 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion:105 Resp: 218603
Ion Ratio Lower Upper
105 100
120 44.3 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 21.39 ug/l
 RT: 11.927 min Scan# 1778
 Delta R.T. -0.000 min
 Lab File: VX020426.D
 Acq: 29 Dec 2020 11:59




Instrument :
 MSVOA_X
 ClientSampleId :
 VX1229WBSD01


Tgt Ion:105 Resp: 240956
 Ion Ratio Lower Upper
 105 100
 134 19.7 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 21.44 ug/l
 RT: 12.049 min Scan# 1798
 Delta R.T. 0.001 min
 Lab File: VX020426.D
 Acq: 29 Dec 2020 11:59

Tgt Ion:119 Resp: 222795
 Ion Ratio Lower Upper
 119 100
 134 26.0 12.9 38.7
 91 23.6 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 20.54 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion:146 Resp: 116510
Ion Ratio Lower Upper
146 100
111 40.9 20.2 60.5
148 63.0 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 19.69 ug/l
RT: 12.079 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion:146 Resp: 116499
Ion Ratio Lower Upper
146 100
111 41.3 19.6 58.8
148 64.9 31.6 95.0









#89
n-Butylbenzene
Concen: 20.97 ug/l
RT: 12.372 min Scan# 1851
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion: 91 Resp: 190870
Ion Ratio Lower Upper
91 100
92 54.4 27.4 82.0
134 25.8 13.2 39.5



#90
Hexachloroethane
Concen: 20.52 ug/l
RT: 12.579 min Scan# 1885
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 117 Resp: 36777
Ion Ratio Lower Upper
117 100
201 84.8 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 20.63 ug/l
RT: 12.372 min Scan# 1851
Delta R.T. -0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion:146 Resp: 116457
Ion Ratio Lower Upper
146 100
111 41.6 20.8 62.2
148 63.7 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 20.73 ug/l
RT: 12.981 min Scan# 1951
Delta R.T. 0.000 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion: 75 Resp: 18963
Ion Ratio Lower Upper
75 100
155 90.3 45.3 135.9
157 115.8 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 21.50 ug/l
RT: 13.628 min Scan# 2057
Delta R.T. 0.001 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59




Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01


Tgt Ion:180 Resp: 78040
Ion Ratio Lower Upper
180 100
182 94.7 47.7 143.1
145 29.9 15.0 45.0



#94
Hexachlorobutadiene
Concen: 25.14 ug/l
RT: 13.762 min Scan# 2079
Delta R.T. -0.006 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion:225 Resp: 36185
Ion Ratio Lower Upper
225 100
223 63.2 31.1 93.2
227 62.5 31.6 95.0









#95
Naphthalene
Concen: 21.71 ug/l
RT: 13.817 min Scan# 2088
Delta R.T. 0.001 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Instrument :
MSVOA_X
ClientSampleId :
VX1229WBSD01

Tgt Ion:128 Resp: 249477
Ion Ratio Lower Upper
128 100
127 12.7 10.3 15.5
129 10.7 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 21.39 ug/l
RT: 14.000 min Scan# 2118
Delta R.T. 0.001 min
Lab File: VX020426.D
Acq: 29 Dec 2020 11:59

Tgt Ion:180 Resp: 76796
Ion Ratio Lower Upper
180 100
182 95.8 47.6 142.9
145 32.6 16.1 48.3

