

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX122920\
 Data File : VX020423.D
 Acq On : 29 Dec 2020 09:57
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 30 02:33:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	116	0.00
2 T	Dichlorodifluoromethane	50.000	44.426	11.1	110	0.00
3 P	Chloromethane	50.000	41.666	16.7	98	0.00
4 C	Vinyl Chloride	50.000	42.777	14.4#	106	0.00
5 T	Bromomethane	50.000	41.523	17.0	102	0.00
6 T	Chloroethane	50.000	44.531	10.9	103	0.00
7 T	Trichlorofluoromethane	50.000	44.803	10.4	114	0.00
8 T	Diethyl Ether	50.000	43.665	12.7	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	44.762	10.5	114	0.00
10 T	Methyl Iodide	50.000	39.573	20.9	92	0.00
11 T	Tert butyl alcohol	250.000	231.515	7.4	112	0.00
12 CM	1,1-Dichloroethene	50.000	44.347	11.3#	111	0.00
13 T	Acrolein	250.000	225.393	9.8	104	0.00
14 T	Allyl chloride	50.000	47.711	4.6	109	0.00
15 T	Acrylonitrile	250.000	245.399	1.8	107	0.00
16 T	Acetone	250.000	252.362	-0.9	117	0.00
17 T	Carbon Disulfide	50.000	43.207	13.6	108	0.00
18 T	Methyl Acetate	50.000	51.251	-2.5	113	0.00
19 T	Methyl tert-butyl Ether	50.000	46.864	6.3	100	0.00
20 T	Methylene Chloride	50.000	43.230	13.5	99	0.00
21 T	trans-1,2-Dichloroethene	50.000	44.496	11.0	105	0.00
22 T	Diisopropyl ether	50.000	47.303	5.4	102	0.00
23 T	Vinyl Acetate	250.000	237.644	4.9	99	0.00
24 P	1,1-Dichloroethane	50.000	46.112	7.8	105	0.00
25 T	2-Butanone	250.000	257.681	-3.1	113	0.00
26 T	2,2-Dichloropropane	50.000	47.215	5.6	112	0.00
27 T	cis-1,2-Dichloroethene	50.000	45.642	8.7	102	0.00
28 T	Bromochloromethane	50.000	44.514	11.0	100	0.00
29 T	Tetrahydrofuran	250.000	243.368	2.7	105	0.00
30 C	Chloroform	50.000	45.304	9.4#	101	0.00
31 T	Cyclohexane	50.000	47.857	4.3	120	0.00
32 T	1,1,1-Trichloroethane	50.000	47.090	5.8	112	0.00
33 S	1,2-Dichloroethane-d4	50.000	41.144	17.7	92	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	103	0.00
35 S	Dibromofluoromethane	50.000	48.061	3.9	99	0.00
36 T	1,1-Dichloropropene	50.000	49.236	1.5	112	0.00
37 T	Ethyl Acetate	50.000	51.673	-3.3	103	0.00
38 T	Carbon Tetrachloride	50.000	51.377	-2.8	114	0.00
39 T	Methylcyclohexane	50.000	52.987	-6.0	124	0.00
40 TM	Benzene	50.000	49.502	1.0	103	0.00
41 T	Methacrylonitrile	50.000	53.089	-6.2	102	0.00
42 TM	1,2-Dichloroethane	50.000	46.660	6.7	93	0.00
43 T	Isopropyl Acetate	50.000	51.872	-3.7	102	0.00
44 TM	Trichloroethene	50.000	49.177	1.6	107	0.00
45 C	1,2-Dichloropropane	50.000	51.080	-2.2#	103	0.00
46 T	Dibromomethane	50.000	47.973	4.1	97	0.00
47 T	Bromodichloromethane	50.000	51.303	-2.6	101	0.00
48 T	Methyl methacrylate	50.000	52.018	-4.0	101	0.00
49 T	1,4-Dioxane	1000.000	1066.118	-6.6	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122920\
 Data File : VX020423.D
 Acq On : 29 Dec 2020 09:57
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 30 02:33:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	47.797	4.4	100	0.00
51 T	4-Methyl-2-Pentanone	250.000	267.247	-6.9	104	0.00
52 CM	Toluene	50.000	50.950	-1.9#	105	0.00
53 T	t-1,3-Dichloropropene	50.000	52.142	-4.3	101	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.269	-4.5	101	0.00
55 T	1,1,2-Trichloroethane	50.000	50.184	-0.4	100	0.00
56 T	Ethyl methacrylate	50.000	54.434	-8.9	103	0.00
57 T	1,3-Dichloropropane	50.000	50.233	-0.5	99	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	253.101	-1.2	99	0.00
59 T	2-Hexanone	250.000	274.128	-9.7	107	0.00
60 T	Dibromochloromethane	50.000	53.109	-6.2	103	0.00
61 T	1,2-Dibromoethane	50.000	50.420	-0.8	99	0.00
62 S	4-Bromofluorobenzene	50.000	48.771	2.5	102	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	100	0.00
64 T	Tetrachloroethene	50.000	52.409	-4.8	117	0.00
65 PM	Chlorobenzene	50.000	49.275	1.5	101	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.482	-5.0	105	0.00
67 C	Ethyl Benzene	50.000	51.536	-3.1#	106	0.00
68 T	m/p-Xylenes	100.000	104.540	-4.5	105	0.00
69 T	o-Xylene	50.000	51.849	-3.7	103	0.00
70 T	Styrene	50.000	53.269	-6.5	103	0.00
71 P	Bromoform	50.000	54.630	-9.3	109	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	101	0.00
73 T	Isopropylbenzene	50.000	53.057	-6.1	111	0.00
74 T	N-aryl acetate	50.000	52.212	-4.4	100	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.040	-0.1	102	0.00
76 T	1,2,3-Trichloropropane	50.000	50.971	-1.9	104	0.00
77 T	Bromobenzene	50.000	49.379	1.2	104	0.00
78 T	n-propylbenzene	50.000	52.740	-5.5	111	0.00
79 T	2-Chlorotoluene	50.000	50.798	-1.6	105	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.647	-7.3	110	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	54.176	-8.4	105	0.00
82 T	4-Chlorotoluene	50.000	50.867	-1.7	104	0.00
83 T	tert-Butylbenzene	50.000	53.271	-6.5	112	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.828	-5.7	105	0.00
85 T	sec-Butylbenzene	50.000	54.346	-8.7	118	0.00
86 T	p-Isopropyltoluene	50.000	54.555	-9.1	114	0.00
87 T	1,3-Dichlorobenzene	50.000	49.808	0.4	102	0.00
88 T	1,4-Dichlorobenzene	50.000	47.731	4.5	101	0.00
89 T	n-Butylbenzene	50.000	54.665	-9.3	117	0.00
90 T	Hexachloroethane	50.000	53.525	-7.0	115	0.00
91 T	1,2-Dichlorobenzene	50.000	50.116	-0.2	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	52.325	-4.7	102	0.00
93 T	1,2,4-Trichlorobenzene	50.000	53.791	-7.6	112	0.00
94 T	Hexachlorobutadiene	50.000	61.694	-23.4	151	0.00
95 T	Naphthalene	50.000	55.745	-11.5	107	0.00
96 T	1,2,3-Trichlorobenzene	50.000	53.667	-7.3	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122920\
Data File : VX020423.D
Acq On : 29 Dec 2020 09:57
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 30 02:33:35 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6