

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122920\
 Data File : VX020445.D
 Acq On : 29 Dec 2020 19:27
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 30 02:40:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	103	0.01
2 T	Dichlorodifluoromethane	50.000	43.324	13.4	95	0.00
3 P	Chloromethane	50.000	44.825	10.3	93	0.00
4 C	Vinyl Chloride	50.000	43.289	13.4#	95	0.00
5 T	Bromomethane	50.000	43.035	13.9	93	0.00
6 T	Chloroethane	50.000	44.062	11.9	90	0.00
7 T	Trichlorofluoromethane	50.000	43.390	13.2	97	0.00
8 T	Diethyl Ether	50.000	43.689	12.6	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	43.709	12.6	98	0.00
10 T	Methyl Iodide	50.000	45.831	8.3	95	0.00
11 T	Tert butyl alcohol	250.000	251.430	-0.6	107	0.00
12 CM	1,1-Dichloroethene	50.000	43.784	12.4#	96	0.00
13 T	Acrolein	250.000	212.485	15.0	86	0.00
14 T	Allyl chloride	50.000	46.803	6.4	95	0.00
15 T	Acrylonitrile	250.000	259.815	-3.9	100	0.00
16 T	Acetone	250.000	217.216	13.1	89	0.00
17 T	Carbon Disulfide	50.000	42.671	14.7	94	0.00
18 T	Methyl Acetate	50.000	53.811	-7.6	105	0.00
19 T	Methyl tert-butyl Ether	50.000	47.985	4.0	90	0.00
20 T	Methylene Chloride	50.000	44.222	11.6	89	0.00
21 T	trans-1,2-Dichloroethene	50.000	44.795	10.4	93	0.00
22 T	Diisopropyl ether	50.000	48.452	3.1	92	0.00
23 T	Vinyl Acetate	250.000	243.337	2.7	89	0.00
24 P	1,1-Dichloroethane	50.000	47.378	5.2	95	0.00
25 T	2-Butanone	250.000	255.368	-2.1	99	0.01
26 T	2,2-Dichloropropane	50.000	41.981	16.0	88	0.00
27 T	cis-1,2-Dichloroethene	50.000	46.800	6.4	92	0.00
28 T	Bromochloromethane	50.000	46.261	7.5	92	0.00
29 T	Tetrahydrofuran	250.000	267.395	-7.0	102	0.00
30 C	Chloroform	50.000	47.907	4.2#	94	0.00
31 T	Cyclohexane	50.000	48.210	3.6	107	0.00
32 T	1,1,1-Trichloroethane	50.000	48.729	2.5	102	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.415	7.2	91	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	97	0.00
35 S	Dibromofluoromethane	50.000	50.243	-0.5	97	0.00
36 T	1,1-Dichloropropene	50.000	47.343	5.3	101	0.00
37 T	Ethyl Acetate	50.000	51.415	-2.8	97	0.00
38 T	Carbon Tetrachloride	50.000	50.483	-1.0	105	0.00
39 T	Methylcyclohexane	50.000	47.301	5.4	104	0.00
40 TM	Benzene	50.000	48.839	2.3	95	0.00
41 T	Methacrylonitrile	50.000	53.578	-7.2	97	0.00
42 TM	1,2-Dichloroethane	50.000	46.855	6.3	88	0.00
43 T	Isopropyl Acetate	50.000	51.306	-2.6	95	0.00
44 TM	Trichloroethene	50.000	47.277	5.4	97	0.00
45 C	1,2-Dichloropropane	50.000	50.403	-0.8#	96	0.00
46 T	Dibromomethane	50.000	47.861	4.3	91	0.00
47 T	Bromodichloromethane	50.000	50.159	-0.3	93	0.00
48 T	Methyl methacrylate	50.000	51.834	-3.7	94	0.00
49 T	1,4-Dioxane	1000.000	1113.108	-11.3	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122920\
 Data File : VX020445.D
 Acq On : 29 Dec 2020 19:27
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 30 02:40:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
50 S	Toluene-d8	50.000	49.453	1.1	97	0.00
51 T	4-Methyl-2-Pentanone	250.000	273.054	-9.2	100	0.00
52 CM	Toluene	50.000	50.058	-0.1#	97	0.00
53 T	t-1,3-Dichloropropene	50.000	50.003	-0.0	91	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.021	-0.0	91	0.00
55 T	1,1,2-Trichloroethane	50.000	50.391	-0.8	94	0.00
56 T	Ethyl methacrylate	50.000	53.257	-6.5	95	0.00
57 T	1,3-Dichloropropane	50.000	50.267	-0.5	93	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	243.029	2.8	89	0.00
59 T	2-Hexanone	250.000	275.150	-10.1	101	0.00
60 T	Dibromochloromethane	50.000	51.595	-3.2	94	0.00
61 T	1,2-Dibromoethane	50.000	49.800	0.4	92	0.00
62 S	4-Bromofluorobenzene	50.000	50.337	-0.7	99	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	94	0.00
64 T	Tetrachloroethene	50.000	49.065	1.9	104	0.00
65 PM	Chlorobenzene	50.000	48.625	2.8	94	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.676	-3.4	98	0.00
67 C	Ethyl Benzene	50.000	49.819	0.4#	97	0.00
68 T	m/p-Xylenes	100.000	99.224	0.8	95	0.00
69 T	o-Xylene	50.000	50.301	-0.6	94	0.00
70 T	Styrene	50.000	51.911	-3.8	95	0.00
71 P	Bromoform	50.000	55.019	-10.0	103	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	98	0.00
73 T	Isopropylbenzene	50.000	49.585	0.8	101	0.00
74 T	N-aryl acetate	50.000	50.231	-0.5	93	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.604	0.8	98	0.00
76 T	1,2,3-Trichloropropane	50.000	49.518	1.0	98	0.00
77 T	Bromobenzene	50.000	46.916	6.2	96	0.00
78 T	n-propylbenzene	50.000	48.667	2.7	99	0.00
79 T	2-Chlorotoluene	50.000	47.729	4.5	95	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.657	0.7	98	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.862	0.3	94	0.00
82 T	4-Chlorotoluene	50.000	48.094	3.8	95	0.00
83 T	tert-Butylbenzene	50.000	50.195	-0.4	102	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.011	2.0	94	0.00
85 T	sec-Butylbenzene	50.000	49.932	0.1	105	0.00
86 T	p-Isopropyltoluene	50.000	50.040	-0.1	101	0.00
87 T	1,3-Dichlorobenzene	50.000	46.264	7.5	92	0.00
88 T	1,4-Dichlorobenzene	50.000	45.548	8.9	94	0.00
89 T	n-Butylbenzene	50.000	48.932	2.1	101	0.00
90 T	Hexachloroethane	50.000	50.060	-0.1	105	0.00
91 T	1,2-Dichlorobenzene	50.000	47.807	4.4	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.710	-1.4	96	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.934	2.1	99	0.00
94 T	Hexachlorobutadiene	50.000	52.781	-5.6	125	0.00
95 T	Naphthalene	50.000	52.595	-5.2	98	0.00
96 T	1,2,3-Trichlorobenzene	50.000	49.509	1.0	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122920\
Data File : VX020445.D
Acq On : 29 Dec 2020 19:27
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 30 02:40:56 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820w.M
Quant Title : Sw846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6