


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122920\
 Data File : VX020422.D
 Acq On : 29 Dec 2020 08:47
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Title : SW846 8260
 Last Update : Fri Dec 18 13:09:49 2020



AutoFind: Scans 1644, 1645, 1646; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	10312	PASS
75	95	30	60	47.9	30851	PASS
95	95	100	100	100.0	64403	PASS
96	95	5	9	6.8	4400	PASS
173	174	0.00	2	0.9	431	PASS
174	95	50	100	72.4	46600	PASS
175	174	5	9	7.2	3349	PASS
176	174	95	101	97.3	45325	PASS
177	176	5	9	6.6	3003	PASS