

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122921\
 Data File : VX026137.D
 Acq On : 29 Dec 2021 14:34
 Operator : JC/MD
 Sample : M5216-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VHBLK001

Quant Time: Dec 30 02:41:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML122821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 30 02:39:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	174051	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	151116	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	69445	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.374	65	74377	55.057	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	110.120%
7) Chloroethane-d5	1.673	69	59237	60.320	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	120.640%
11) 1,1-Dichloroethene-d2	2.313	63	108945	44.182	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.360%
21) 2-Butanone-d5	4.459	46	117043	125.386	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	125.390%
24) Chloroform-d	5.068	84	131226	57.199	ug/L	0.01
Spiked Amount	50.000	Range	70 - 125	Recovery	=	114.400%
26) 1,2-Dichloroethane-d4	5.964	65	90570	60.902	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	121.800%
32) Benzene-d6	5.983	84	260969	59.900	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	119.800%
36) 1,2-Dichloropropane-d6	7.318	67	83274	61.786	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	123.580%#
41) Toluene-d8	8.653	98	247265	62.231	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	124.460%#
43) trans-1,3-Dichloroprop...	8.952	79	40700	61.862	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	123.720%
47) 2-Hexanone-d5	9.385	63	82327	130.973	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	130.970%#
56) 1,1,2,2-Tetrachloroeth...	11.195	84	110415	61.555	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	123.120%#
66) 1,2-Dichlorobenzene-d4	12.323	152	82400	62.203	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	124.400%#
Target Compounds					Qvalue	
13) Acetone	2.392	43	2167	2.451	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122921\
Data File : VX026137.D
Acq On : 29 Dec 2021 14:34
Operator : JC/MD
Sample : M5216-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VHBLK001

Quant Time: Dec 30 02:41:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM122821WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 30 02:39:24 2021
Response via : Initial Calibration

