

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122921\
 Data File : VX026139.D
 Acq On : 29 Dec 2021 15:20
 Operator : JC/MD
 Sample : VIBLK623
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK623

Quant Time: Dec 30 02:42:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML122821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 30 02:39:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	190202	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	165310	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	76490	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.374	65	72642	49.207	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.420%
7) Chloroethane-d5	1.672	69	57552	53.628	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	107.260%
11) 1,1-Dichloroethene-d2	2.313	63	106205	39.413	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.820%
21) 2-Butanone-d5	4.459	46	112728	110.508	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	110.510%
24) Chloroform-d	5.068	84	122754	48.963	ug/L	0.01
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.920%
26) 1,2-Dichloroethane-d4	5.964	65	87799	54.025	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.060%
32) Benzene-d6	5.983	84	253277	53.143	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.280%
36) 1,2-Dichloropropane-d6	7.312	67	78184	53.029	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.060%
41) Toluene-d8	8.653	98	234120	53.863	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.720%
43) trans-1,3-Dichloroprop...	8.952	79	38410	53.368	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	106.740%
47) 2-Hexanone-d5	9.384	63	78893	114.733	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	114.730%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	104791	53.404	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	106.800%
66) 1,2-Dichlorobenzene-d4	12.323	152	81539	55.884	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122921\
Data File : VX026139.D
Acq On : 29 Dec 2021 15:20
Operator : JC/MD
Sample : VIBLK623
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK623

Quant Time: Dec 30 02:42:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM122821WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 30 02:39:24 2021
Response via : Initial Calibration

