

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122921\
 Data File : VX026140.D
 Acq On : 29 Dec 2021 15:44
 Operator : JC/MD
 Sample : M5216-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COHB6

Quant Time: Dec 30 02:42:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML122821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 30 02:39:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	192317	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	167199	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	77860	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.374	65	74661	50.018	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 100.040%	
7) Chloroethane-d5	1.672	69	59971	55.267	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 110.540%	
11) 1,1-Dichloroethene-d2	2.312	63	109104	40.044	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 80.080%	
21) 2-Butanone-d5	4.458	46	116518	112.968	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 112.970%	
24) Chloroform-d	5.068	84	131401	51.836	ug/L	0.01
Spiked Amount	50.000	Range	70 - 125	Recovery	= 103.680%	
26) 1,2-Dichloroethane-d4	5.964	65	90834	55.278	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 110.560%	
32) Benzene-d6	5.983	84	262710	54.499	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 109.000%	
36) 1,2-Dichloropropane-d6	7.312	67	81914	54.931	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 109.860%	
41) Toluene-d8	8.653	98	239882	54.565	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 109.140%	
43) trans-1,3-Dichloroprop...	8.952	79	38928	53.477	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 106.960%	
47) 2-Hexanone-d5	9.384	63	78743	113.221	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 113.220%	
56) 1,1,2,2-Tetrachloroeth...	11.195	84	107226	54.027	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 108.060%	
66) 1,2-Dichlorobenzene-d4	12.323	152	85255	57.402	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 114.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122921\
Data File : VX026140.D
Acq On : 29 Dec 2021 15:44
Operator : JC/MD
Sample : M5216-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COHB6

Quant Time: Dec 30 02:42:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM122821WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 30 02:39:24 2021
Response via : Initial Calibration

