

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122921\
 Data File : VX026142.D
 Acq On : 29 Dec 2021 16:30
 Operator : JC/MD
 Sample : M5233-06
 Misc : 8.49g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GBCC4

Quant Time: Dec 30 02:42:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML122821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 30 02:39:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	270023	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.061	117	231430	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	113683	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.374	65	89529	42.719	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.440%
7) Chloroethane-d5	1.660	69	53722	35.261	ug/L	-0.01
Spiked Amount	50.000	Range	70 - 130	Recovery	=	70.520%
11) 1,1-Dichloroethene-d2	2.288	63	144644	37.811	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.620%
21) 2-Butanone-d5	4.483	46	140347	96.913	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.910%
24) Chloroform-d	5.062	84	150809	42.372	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.740%
26) 1,2-Dichloroethane-d4	5.964	65	114740	49.732	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.460%
32) Benzene-d6	5.977	84	341596	51.196	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.400%
36) 1,2-Dichloropropane-d6	7.312	67	104979	50.860	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.720%
41) Toluene-d8	8.653	98	299688	49.250	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.500%
43) trans-1,3-Dichloroprop...	8.952	79	46371	46.022	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.040%
47) 2-Hexanone-d5	9.403	63	97134	100.903	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.900%
56) 1,1,2,2-Tetrachloroeth...	11.201	84	140404	51.110	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.220%
66) 1,2-Dichlorobenzene-d4	12.323	152	114062	52.598	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.200%
Target Compounds						
25) Chloroform	5.105	83	21138	6.445	ug/L	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122921\
Data File : VX026142.D
Acq On : 29 Dec 2021 16:30
Operator : JC/MD
Sample : M5233-06
Misc : 8.49g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GBCC4

Quant Time: Dec 30 02:42:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM122821WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 30 02:39:24 2021
Response via : Initial Calibration

