

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122923\
 Data File : VX039532.D
 Acq On : 29 Dec 2023 10:06
 Operator : JC/MD
 Sample : VX1229WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1229WBL01

Quant Time: Dec 30 01:53:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	82533	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	176716	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	168440	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69948	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72524	47.967	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.940%
35) Dibromofluoromethane	5.385	113	57845	47.381	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.760%
50) Toluene-d8	8.647	98	219020	48.323	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.640%
62) 4-Bromofluorobenzene	11.079	95	84978	45.442	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122923\
Data File : VX039532.D
Acq On : 29 Dec 2023 10:06
Operator : JC/MD
Sample : VX1229WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1229WBL01

Quant Time: Dec 30 01:53:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

