

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122923\
 Data File : VX039543.D
 Acq On : 29 Dec 2023 14:22
 Operator : JC/MD
 Sample : 06031-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MR-321-10-12

Quant Time: Dec 30 02:00:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	83991	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	176875	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	181039	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76787	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72994	47.440	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.880%
35) Dibromofluoromethane	5.385	113	57304	46.896	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.800%
50) Toluene-d8	8.647	98	230614	50.835	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.680%
62) 4-Bromofluorobenzene	11.079	95	93737	50.080	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.160%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	48092	33.486	ug/l	94
20) Methylene Chloride	2.788	84	6416	5.002	ug/l	94
25) 2-Butanone	4.581	43	10814	8.145	ug/l #	87
29) Tetrahydrofuran	5.026	42	5568	7.594	ug/l	93
43) Isopropyl Acetate	6.342	43	94043	24.276	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122923\
Data File : VX039543.D
Acq On : 29 Dec 2023 14:22
Operator : JC/MD
Sample : 06031-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MR-321-10-12

Quant Time: Dec 30 02:00:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

