

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122923\
 Data File : VX039548.D
 Acq On : 29 Dec 2023 16:17
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Dec 30 03:35:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.556	168	85686	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	180874	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	173495	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68317	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75607	48.166	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.340%
35) Dibromofluoromethane	5.385	113	57333	45.882	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.760%
50) Toluene-d8	8.647	98	227620	49.066	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.140%
62) 4-Bromofluorobenzene	11.079	95	86288	45.081	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.160%
Target Compounds						
95) Naphthalene	13.774	128	48762	9.760	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122923\
 Data File : VX039548.D
 Acq On : 29 Dec 2023 16:17
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Dec 30 03:35:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

