

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122923\
 Data File : VX039550.D
 Acq On : 29 Dec 2023 17:02
 Operator : JC/MD
 Sample : 06046-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Dec 30 02:04:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	85197	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	181431	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	184999	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80191	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	75419	48.322	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.640%
35) Dibromofluoromethane	5.385	113	58977	47.053	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.100%
50) Toluene-d8	8.647	98	231621	49.775	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.560%
62) 4-Bromofluorobenzene	11.079	95	95484	49.733	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.460%
Target Compounds						
95) Naphthalene	13.774	128	20918	3.567	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122923\
Data File : VX039550.D
Acq On : 29 Dec 2023 17:02
Operator : JC/MD
Sample : 06046-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Dec 30 02:04:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

