

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122923\
 Data File : VX039551.D
 Acq On : 29 Dec 2023 17:25
 Operator : JC/MD
 Sample : 06046-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 30 02:04:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	81316	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	172878	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	174028	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77701	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	73430	49.293	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.580%
35) Dibromofluoromethane	5.391	113	56156	47.019	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.040%
50) Toluene-d8	8.647	98	221028	49.849	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.700%
62) 4-Bromofluorobenzene	11.079	95	91070	49.780	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.560%
Target Compounds						
95) Naphthalene	13.774	128	16720	2.942	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122923\
Data File : VX039551.D
Acq On : 29 Dec 2023 17:25
Operator : JC/MD
Sample : 06046-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 30 02:04:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

