

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX123019\
 Data File : VX014365.D
 Acq On : 30 Dec 2019 19:40
 Operator : JC/SP
 Sample : K6114-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EO-O1-122719

Quant Time: Dec 31 03:45:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	495813	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	779158	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	705712	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	304514	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	281999	50.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.36%	
35) Dibromofluoromethane	5.49	113	233325	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
50) Toluene-d8	8.71	98	936086	50.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.52%	
62) 4-Bromofluorobenzene	11.14	95	314006	46.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.16%	
Target Compounds						
16) Acetone	2.43	43	89742	24.599	ug/l	95
20) Methylene Chloride	2.85	84	11632	2.025	ug/l	93
43) Isopropyl Acetate	6.43	43	121741	9.318	ug/l	99
95) Naphthalene	13.83	128	50369	2.534	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX123019\
Data File : VX014365.D
Acq On : 30 Dec 2019 19:40
Operator : JC/SP
Sample : K6114-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EO-O1-122719

Quant Time: Dec 31 03:45:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

