

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX123020\
 Data File : VX020447.D
 Acq On : 30 Dec 2020 11:42
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 31 01:03:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	118	0.00
2 T	Dichlorodifluoromethane	50.000	45.302	9.4	114	0.00
3 P	Chloromethane	50.000	44.697	10.6	107	0.00
4 C	Vinyl Chloride	50.000	44.467	11.1#	112	0.00
5 T	Bromomethane	50.000	39.017	22.0	97	0.00
6 T	Chloroethane	50.000	43.203	13.6	102	0.00
7 T	Trichlorofluoromethane	50.000	43.418	13.2	112	0.00
8 T	Diethyl Ether	50.000	42.187	15.6	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.471	3.1	126	0.00
10 T	Methyl Iodide	50.000	43.451	13.1	104	0.00
11 T	Tert butyl alcohol	250.000	217.080	13.2	106	0.00
12 CM	1,1-Dichloroethene	50.000	46.213	7.6#	117	0.00
13 T	Acrolein	250.000	224.605	10.2	105	0.00
14 T	Allyl chloride	50.000	45.355	9.3	106	0.00
15 T	Acrylonitrile	250.000	238.687	4.5	106	0.00
16 T	Acetone	250.000	243.322	2.7	114	0.00
17 T	Carbon Disulfide	50.000	41.364	17.3	105	0.00
18 T	Methyl Acetate	50.000	50.599	-1.2	114	0.00
19 T	Methyl tert-butyl Ether	50.000	44.844	10.3	97	0.00
20 T	Methylene Chloride	50.000	43.987	12.0	102	0.00
21 T	trans-1,2-Dichloroethene	50.000	42.853	14.3	103	0.00
22 T	Diisopropyl ether	50.000	46.651	6.7	102	0.00
23 T	Vinyl Acetate	250.000	231.825	7.3	98	0.00
24 P	1,1-Dichloroethane	50.000	45.141	9.7	104	0.00
25 T	2-Butanone	250.000	249.512	0.2	111	0.00
26 T	2,2-Dichloropropane	50.000	45.599	8.8	109	0.00
27 T	cis-1,2-Dichloroethene	50.000	44.448	11.1	101	0.00
28 T	Bromochloromethane	50.000	46.168	7.7	106	0.00
29 T	Tetrahydrofuran	250.000	235.174	5.9	103	0.00
30 C	Chloroform	50.000	44.439	11.1#	100	0.00
31 T	Cyclohexane	50.000	45.267	9.5	115	0.00
32 T	1,1,1-Trichloroethane	50.000	45.427	9.1	110	0.00
33 S	1,2-Dichloroethane-d4	50.000	45.027	9.9	102	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	108	0.00
35 S	Dibromofluoromethane	50.000	49.379	1.2	107	0.00
36 T	1,1-Dichloropropene	50.000	45.939	8.1	110	0.00
37 T	Ethyl Acetate	50.000	49.645	0.7	105	0.00
38 T	Carbon Tetrachloride	50.000	48.197	3.6	113	0.00
39 T	Methylcyclohexane	50.000	48.772	2.5	120	0.00
40 TM	Benzene	50.000	46.882	6.2	102	0.00
41 T	Methacrylonitrile	50.000	49.390	1.2	100	0.00
42 TM	1,2-Dichloroethane	50.000	45.126	9.7	95	0.00
43 T	Isopropyl Acetate	50.000	47.441	5.1	99	0.00
44 TM	Trichloroethene	50.000	47.427	5.1	109	0.00
45 C	1,2-Dichloropropane	50.000	49.496	1.0#	105	0.00
46 T	Dibromomethane	50.000	45.570	8.9	97	0.00
47 T	Bromodichloromethane	50.000	47.612	4.8	99	0.00
48 T	Methyl methacrylate	50.000	46.791	6.4	95	0.00
49 T	1,4-Dioxane	1000.000	980.733	1.9	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123020\
 Data File : VX020447.D
 Acq On : 30 Dec 2020 11:42
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 31 01:03:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	50.343	-0.7	111	0.00
51 T	4-Methyl-2-Pentanone	250.000	253.200	-1.3	104	0.00
52 CM	Toluene	50.000	48.771	2.5#	105	0.00
53 T	t-1,3-Dichloropropene	50.000	48.805	2.4	99	0.00
54 T	cis-1,3-Dichloropropene	50.000	48.860	2.3	100	0.00
55 T	1,1,2-Trichloroethane	50.000	49.765	0.5	104	0.00
56 T	Ethyl methacrylate	50.000	50.351	-0.7	100	0.00
57 T	1,3-Dichloropropane	50.000	49.041	1.9	102	0.00
58 T	2-Chloroethyl vinyl ether	250.000	230.068	8.0	95	0.00
59 T	2-Hexanone	250.000	267.484	-7.0	110	0.00
60 T	Dibromochloromethane	50.000	51.607	-3.2	105	0.00
61 T	1,2-Dibromoethane	50.000	48.816	2.4	101	0.00
62 S	4-Bromofluorobenzene	50.000	49.916	0.2	110	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	105	0.00
64 T	Tetrachloroethene	50.000	50.857	-1.7	120	0.00
65 PM	Chlorobenzene	50.000	47.033	5.9	102	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	48.755	2.5	103	0.00
67 C	Ethyl Benzene	50.000	47.995	4.0#	104	0.00
68 T	m/p-Xylenes	100.000	96.967	3.0	103	0.00
69 T	o-Xylene	50.000	48.426	3.1	101	0.00
70 T	Styrene	50.000	49.888	0.2	101	0.00
71 P	Bromoform	50.000	51.321	-2.6	107	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	112	0.00
73 T	Isopropylbenzene	50.000	46.004	8.0	108	0.00
74 T	N-aryl acetate	50.000	44.128	11.7	94	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	44.698	10.6	101	0.00
76 T	1,2,3-Trichloropropane	50.000	44.958	10.1	103	0.00
77 T	Bromobenzene	50.000	44.282	11.4	104	0.00
78 T	n-propylbenzene	50.000	47.336	5.3	111	0.00
79 T	2-Chlorotoluene	50.000	45.390	9.2	104	0.00
80 T	1,3,5-Trimethylbenzene	50.000	47.552	4.9	108	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.115	5.8	102	0.00
82 T	4-Chlorotoluene	50.000	45.876	8.2	105	0.00
83 T	tert-Butylbenzene	50.000	46.662	6.7	109	0.00
84 T	1,2,4-Trimethylbenzene	50.000	47.946	4.1	106	0.00
85 T	sec-Butylbenzene	50.000	49.906	0.2	121	0.00
86 T	p-Isopropyltoluene	50.000	49.661	0.7	116	0.00
87 T	1,3-Dichlorobenzene	50.000	46.717	6.6	107	0.00
88 T	1,4-Dichlorobenzene	50.000	45.323	9.4	107	0.00
89 T	n-Butylbenzene	50.000	50.404	-0.8	120	0.00
90 T	Hexachloroethane	50.000	48.898	2.2	118	0.00
91 T	1,2-Dichlorobenzene	50.000	46.240	7.5	107	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	47.920	4.2	105	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.409	3.2	113	0.00
94 T	Hexachlorobutadiene	50.000	54.061	-8.1	147	0.00
95 T	Naphthalene	50.000	50.004	-0.0	107	0.00
96 T	1,2,3-Trichlorobenzene	50.000	48.464	3.1	113	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123020\
Data File : VX020447.D
Acq On : 30 Dec 2020 11:42
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 31 01:03:42 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6