

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX123119\
 Data File : VX014371.D
 Acq On : 31 Dec 2019 11:38
 Operator : JC/SP
 Sample : VX1231WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1231WBL01

Quant Time: Jan 01 03:48:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	522398	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	802895	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	713493	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	302783	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	292603	49.42	ug/l	0.00
Spiked Amount						
			Recovery	=	98.84%	
35) Dibromofluoromethane	5.48	113	243585	49.91	ug/l	0.00
Spiked Amount						
			Recovery	=	99.82%	
50) Toluene-d8	8.71	98	953667	50.19	ug/l	0.00
Spiked Amount						
			Recovery	=	100.38%	
62) 4-Bromofluorobenzene	11.13	95	313196	45.09	ug/l	0.00
Spiked Amount						
			Recovery	=	90.18%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX123119\
Data File : VX014371.D
Acq On : 31 Dec 2019 11:38
Operator : JC/SP
Sample : VX1231WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1231WBL01

Quant Time: Jan 01 03:48:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

