

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX123119\
 Data File : VX014376.D
 Acq On : 31 Dec 2019 13:44
 Operator : JC/SP
 Sample : K6462-01ME 40X
 Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 K500-1ME

Quant Time: Jan 02 01:22:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	529397	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	812254	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	746727	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	352534	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	291791	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
35) Dibromofluoromethane	5.49	113	237808	48.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.32%	
50) Toluene-d8	8.71	98	968595	50.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.78%	
62) 4-Bromofluorobenzene	11.13	95	348412	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
Target Compounds						
68) m/p-Xylenes	10.36	106	78861	7.512	ug/l	100
69) o-Xylene	10.70	106	46932	4.558	ug/l	99
80) 1,3,5-Trimethylbenzene	11.50	105	101912	4.879	ug/l	100
84) 1,2,4-Trimethylbenzene	11.80	105	326818	15.623	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX123119\
Data File : VX014376.D
Acq On : 31 Dec 2019 13:44
Operator : JC/SP
Sample : K6462-01ME 40X
Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
K500-1ME

Quant Time: Jan 02 01:22:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

