

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX123119\
 Data File : VX014370.D
 Acq On : 31 Dec 2019 11:00
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 01 03:42:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	78	0.00
2 T	Dichlorodifluoromethane	0.447	0.366	18.1	67	0.00
3 P	Chloromethane	0.600	0.517	13.8	70	0.00
4 C	Vinyl Chloride	0.633	0.541	14.5#	70	0.00
5 T	Bromomethane	0.450	0.414	8.0	78	0.00
6 T	Chloroethane	0.391	0.359	8.2	76	0.01
7 T	Trichlorofluoromethane	0.784	0.718	8.4	75	0.00
8 T	Diethyl Ether	0.368	0.358	2.7	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.473	0.456	3.6	80	0.00
10 T	Methyl Iodide	0.575	0.503	12.5	64	0.00
11 T	Tert butyl alcohol	0.122	0.120	1.6	84	-0.02
12 CM	1,1-Dichloroethene	0.481	0.433	10.0#	74	0.00
13 T	Acrolein	0.070	0.083	-18.6	101	0.00
14 T	Allyl chloride	0.859	0.865	-0.7	81	0.00
15 T	Acrylonitrile	0.286	0.311	-8.7	89	0.00
16 T	Acetone	0.368	0.295	19.8	63	0.00
17 T	Carbon Disulfide	1.418	1.088	23.3#	66	0.00
18 T	Methyl Acetate	0.729	0.798	-9.5	91	0.00
19 T	Methyl tert-butyl Ether	1.580	1.664	-5.3	85	0.00
20 T	Methylene Chloride	0.579	0.554	4.3	83	0.00
21 T	trans-1,2-Dichloroethene	0.530	0.481	9.2	76	0.00
22 T	Diisopropyl ether	1.711	1.846	-7.9	87	0.00
23 T	Vinyl Acetate	1.397	1.583	-13.3	90	0.00
24 P	1,1-Dichloroethane	0.952	0.941	1.2	82	0.00
25 T	2-Butanone	0.454	0.450	0.9	79	0.00
26 T	2,2-Dichloropropane	0.739	0.717	3.0	78	0.00
27 T	cis-1,2-Dichloroethene	0.599	0.595	0.7	82	0.00
28 T	Bromochloromethane	0.332	0.422	-27.1#	89	0.00
29 T	Tetrahydrofuran	0.254	0.275	-8.3	88	-0.01
30 C	Chloroform	0.903	0.919	-1.8#	82	0.00
31 T	Cyclohexane	0.847	0.802	5.3	75	0.00
32 T	1,1,1-Trichloroethane	0.769	0.739	3.9	78	0.00
33 S	1,2-Dichloroethane-d4	0.567	0.621	-9.5	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	78	0.00
35 S	Dibromofluoromethane	0.304	0.341	-12.2	89	0.00
36 T	1,1-Dichloropropene	0.459	0.436	5.0	77	0.00
37 T	Ethyl Acetate	0.507	0.550	-8.5	86	0.00
38 T	Carbon Tetrachloride	0.418	0.407	2.6	76	0.00
39 T	Methylcyclohexane	0.566	0.536	5.3	77	0.00
40 TM	Benzene	1.412	1.392	1.4	81	0.00
41 T	Methacrylonitrile	0.281	0.307	-9.3	89	0.00
42 TM	1,2-Dichloroethane	0.479	0.490	-2.3	84	0.00
43 T	Isopropyl Acetate	0.838	0.903	-7.8	86	0.00
44 TM	Trichloroethene	0.390	0.375	3.8	81	0.00
45 C	1,2-Dichloropropane	0.362	0.383	-5.8#	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX123119\
 Data File : VX014370.D
 Acq On : 31 Dec 2019 11:00
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 01 03:42:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.239	0.243	-1.7	85	0.00
47 T	Bromodichloromethane	0.451	0.477	-5.8	84	0.00
48 T	Methyl methacrylate	0.411	0.451	-9.7	87	0.00
49 T	1,4-Dioxane	0.008	0.008	0.0	80	0.00
50 S	Toluene-d8	1.183	1.274	-7.7	84	0.00
51 T	4-Methyl-2-Pentanone	0.527	0.570	-8.2	87	0.00
52 CM	Toluene	0.892	0.874	2.0#	80	0.00
53 T	t-1,3-Dichloropropene	0.494	0.536	-8.5	83	0.00
54 T	cis-1,3-Dichloropropene	0.550	0.601	-9.3	84	0.00
55 T	1,1,2-Trichloroethane	0.358	0.383	-7.0	87	0.00
56 T	Ethyl methacrylate	0.562	0.608	-8.2	84	0.00
57 T	1,3-Dichloropropane	0.602	0.635	-5.5	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.213	0.246	-15.5	87	0.00
59 T	2-Hexanone	0.424	0.435	-2.6	80	0.00
60 T	Dibromochloromethane	0.355	0.389	-9.6	84	0.00
61 T	1,2-Dibromoethane	0.366	0.386	-5.5	84	0.00
62 S	4-Bromofluorobenzene	0.433	0.466	-7.6	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	80	0.00
64 T	Tetrachloroethene	0.442	0.407	7.9	77	0.00
65 PM	Chlorobenzene	1.063	1.028	3.3	81	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.385	-1.3	83	0.00
67 C	Ethyl Benzene	1.828	1.796	1.8#	79	0.00
68 T	m/p-Xylenes	0.703	0.680	3.3	79	0.00
69 T	o-Xylene	0.689	0.668	3.0	79	0.00
70 T	Styrene	1.166	1.170	-0.3	80	0.00
71 P	Bromoform	0.305	0.320	-4.9	83	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
73 T	Isopropylbenzene	3.520	3.509	0.3	80	0.00
74 T	N-amyl acetate	1.650	1.694	-2.7	81	0.00
75 P	1,1,2,2-Tetrachloroethane	1.220	1.240	-1.6	82	0.00
76 T	1,2,3-Trichloropropane	1.085	1.024	5.6	74	0.00
77 T	Bromobenzene	0.969	0.928	4.2	82	0.00
78 T	n-propylbenzene	3.949	3.988	-1.0	80	0.00
79 T	2-Chlorotoluene	2.400	2.348	2.2	79	0.00
80 T	1,3,5-Trimethylbenzene	2.963	2.917	1.6	79	0.00
81 T	trans-1,4-Dichloro-2-butene	0.385	0.416	-8.1	82	0.00
82 T	4-Chlorotoluene	2.784	2.738	1.7	81	0.00
83 T	tert-Butylbenzene	2.883	2.793	3.1	79	0.00
84 T	1,2,4-Trimethylbenzene	2.967	2.977	-0.3	80	0.00
85 T	sec-Butylbenzene	3.404	3.397	0.2	80	0.00
86 T	p-Isopropyltoluene	3.114	3.107	0.2	80	0.00
87 T	1,3-Dichlorobenzene	1.698	1.636	3.7	81	0.00
88 T	1,4-Dichlorobenzene	1.726	1.629	5.6	81	0.00
89 T	n-Butylbenzene	2.657	2.748	-3.4	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX123119\
Data File : VX014370.D
Acq On : 31 Dec 2019 11:00
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 01 03:42:02 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.559	0.548	2.0	78	0.00
91 T	1,2-Dichlorobenzene	1.709	1.665	2.6	81	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.270	0.258	4.4	81	0.00
93 T	1,2,4-Trichlorobenzene	1.111	1.121	-0.9	83	0.00
94 T	Hexachlorobutadiene	0.534	0.526	1.5	83	0.00
95 T	Naphthalene	3.263	3.513	-7.7	84	0.00
96 T	1,2,3-Trichlorobenzene	1.096	1.146	-4.6	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6