


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX123119\
 Data File : VX014369.D
 Acq On : 31 Dec 2019 09:53
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Title : SW846 8260
 Last Update : Tue Dec 17 03:01:07 2019



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1670

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	20087	PASS
75	95	30	60	49.0	50192	PASS
95	95	100	100	100.0	102525	PASS
96	95	5	9	7.2	7333	PASS
173	174	0.00	2	1.3	1102	PASS
174	95	50	100	83.0	85145	PASS
175	174	5	9	7.6	6487	PASS
176	174	95	101	97.9	83398	PASS
177	176	5	9	6.6	5475	PASS