

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX123120\
 Data File : VX020470.D
 Acq On : 31 Dec 2020 09:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 01 01:07:26 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	112	0.00
2 T	Dichlorodifluoromethane	50.000	45.926	8.1	110	0.00
3 P	Chloromethane	50.000	41.858	16.3	95	0.00
4 C	Vinyl Chloride	50.000	43.793	12.4#	105	0.00
5 T	Bromomethane	50.000	47.591	4.8	112	0.00
6 T	Chloroethane	50.000	44.991	10.0	100	0.00
7 T	Trichlorofluoromethane	50.000	47.654	4.7	116	0.00
8 T	Diethyl Ether	50.000	44.628	10.7	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.553	2.9	119	0.00
10 T	Methyl Iodide	50.000	44.110	11.8	100	0.00
11 T	Tert butyl alcohol	250.000	224.527	10.2	104	0.00
12 CM	1,1-Dichloroethene	50.000	45.634	8.7#	109	0.00
13 T	Acrolein	250.000	223.509	10.6	99	0.00
14 T	Allyl chloride	50.000	44.975	10.0	99	0.00
15 T	Acrylonitrile	250.000	236.630	5.3	99	0.00
16 T	Acetone	250.000	271.514	-8.6	121	0.00
17 T	Carbon Disulfide	50.000	43.185	13.6	104	0.00
18 T	Methyl Acetate	50.000	50.064	-0.1	106	0.00
19 T	Methyl tert-butyl Ether	50.000	47.789	4.4	98	0.00
20 T	Methylene Chloride	50.000	43.549	12.9	96	0.00
21 T	trans-1,2-Dichloroethene	50.000	44.823	10.4	102	0.00
22 T	Diisopropyl ether	50.000	47.226	5.5	98	0.00
23 T	Vinyl Acetate	250.000	237.125	5.1	95	0.00
24 P	1,1-Dichloroethane	50.000	46.708	6.6	102	0.00
25 T	2-Butanone	250.000	258.161	-3.3	109	0.00
26 T	2,2-Dichloropropane	50.000	48.327	3.3	110	0.00
27 T	cis-1,2-Dichloroethene	50.000	45.579	8.8	98	0.00
28 T	Bromochloromethane	50.000	48.100	3.8	104	0.00
29 T	Tetrahydrofuran	250.000	238.310	4.7	99	0.00
30 C	Chloroform	50.000	47.378	5.2#	101	0.00
31 T	Cyclohexane	50.000	45.714	8.6	110	0.00
32 T	1,1,1-Trichloroethane	50.000	48.996	2.0	112	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.281	3.4	103	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	102	0.00
35 S	Dibromofluoromethane	50.000	53.271	-6.5	108	0.00
36 T	1,1-Dichloropropene	50.000	48.850	2.3	109	0.00
37 T	Ethyl Acetate	50.000	49.931	0.1	99	0.00
38 T	Carbon Tetrachloride	50.000	52.447	-4.9	115	0.00
39 T	Methylcyclohexane	50.000	50.156	-0.3	116	0.00
40 TM	Benzene	50.000	48.408	3.2	99	0.00
41 T	Methacrylonitrile	50.000	50.346	-0.7	96	0.00
42 TM	1,2-Dichloroethane	50.000	49.638	0.7	98	0.00
43 T	Isopropyl Acetate	50.000	49.830	0.3	97	0.00
44 TM	Trichloroethene	50.000	48.686	2.6	105	0.00
45 C	1,2-Dichloropropane	50.000	49.420	1.2#	99	0.00
46 T	Dibromomethane	50.000	48.621	2.8	97	0.00
47 T	Bromodichloromethane	50.000	52.621	-5.2	103	0.00
48 T	Methyl methacrylate	50.000	50.781	-1.6	97	0.00
49 T	1,4-Dioxane	1000.000	1032.559	-3.3	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123120\
 Data File : VX020470.D
 Acq On : 31 Dec 2020 09:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 01 01:07:26 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	51.513	-3.0	106	0.00
51 T	4-Methyl-2-Pentanone	250.000	260.541	-4.2	100	0.00
52 CM	Toluene	50.000	49.764	0.5#	101	0.00
53 T	t-1,3-Dichloropropene	50.000	52.124	-4.2	100	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.457	-2.9	99	0.00
55 T	1,1,2-Trichloroethane	50.000	50.839	-1.7	100	0.00
56 T	Ethyl methacrylate	50.000	51.988	-4.0	97	0.00
57 T	1,3-Dichloropropane	50.000	49.852	0.3	98	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	230.350	7.9	89	0.00
59 T	2-Hexanone	250.000	276.210	-10.5	106	0.00
60 T	Dibromochloromethane	50.000	54.688	-9.4	105	0.00
61 T	1,2-Dibromoethane	50.000	50.472	-0.9	98	0.00
62 S	4-Bromofluorobenzene	50.000	53.626	-7.3	111	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	99	0.00
64 T	Tetrachloroethene	50.000	52.376	-4.8	116	0.00
65 PM	Chlorobenzene	50.000	48.735	2.5	99	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.846	-3.7	103	0.00
67 C	Ethyl Benzene	50.000	51.099	-2.2#	104	0.00
68 T	m/p-Xylenes	100.000	103.825	-3.8	104	0.00
69 T	o-Xylene	50.000	51.061	-2.1	101	0.00
70 T	Styrene	50.000	53.264	-6.5	102	0.00
71 P	Bromoform	50.000	56.712	-13.4	112	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	106	0.00
73 T	Isopropylbenzene	50.000	50.004	-0.0	110	0.00
74 T	N-aryl acetate	50.000	49.365	1.3	99	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.562	0.9	106	0.00
76 T	1,2,3-Trichloropropane	50.000	49.227	1.5	106	0.00
77 T	Bromobenzene	50.000	47.405	5.2	105	0.00
78 T	n-propylbenzene	50.000	51.064	-2.1	113	0.00
79 T	2-Chlorotoluene	50.000	49.183	1.6	106	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.956	0.1	107	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	53.198	-6.4	109	0.00
82 T	4-Chlorotoluene	50.000	48.295	3.4	104	0.00
83 T	tert-Butylbenzene	50.000	49.121	1.8	108	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.395	-0.8	105	0.00
85 T	sec-Butylbenzene	50.000	50.971	-1.9	116	0.00
86 T	p-Isopropyltoluene	50.000	51.095	-2.2	112	0.00
87 T	1,3-Dichlorobenzene	50.000	47.907	4.2	104	0.00
88 T	1,4-Dichlorobenzene	50.000	46.527	6.9	104	0.00
89 T	n-Butylbenzene	50.000	51.479	-3.0	116	0.00
90 T	Hexachloroethane	50.000	52.544	-5.1	119	0.00
91 T	1,2-Dichlorobenzene	50.000	47.794	4.4	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.709	2.6	100	0.00
93 T	1,2,4-Trichlorobenzene	50.000	47.868	4.3	105	0.00
94 T	Hexachlorobutadiene	50.000	56.765	-13.5	146	0.00
95 T	Naphthalene	50.000	50.650	-1.3	102	0.00
96 T	1,2,3-Trichlorobenzene	50.000	49.574	0.9	109	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123120\
Data File : VX020470.D
Acq On : 31 Dec 2020 09:36
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 01 01:07:26 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6