

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123120\
 Data File : VX020493.D
 Acq On : 31 Dec 2020 19:07
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 01 00:14:17 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.01
2 T	Dichlorodifluoromethane	0.543	0.491	9.6	96	0.00
3 P	Chloromethane	0.630	0.556	11.7	89	0.00
4 C	Vinyl Chloride	0.686	0.620	9.6#	96	0.00
5 T	Bromomethane	0.220	0.217	1.4	103	0.00
6 T	Chloroethane	0.427	0.402	5.9	92	0.00
7 T	Trichlorofluoromethane	0.882	0.859	2.6	105	0.00
8 T	Diethyl Ether	0.384	0.370	3.6	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.499	0.479	4.0	104	0.00
10 T	Methyl Iodide	0.622	0.668	-7.4	100	0.00
11 T	Tert butyl alcohol	0.150	0.152	-1.3	103	0.00
12 CM	1,1-Dichloroethene	0.511	0.480	6.1#	100	0.00
13 T	Acrolein	0.094	0.106	-12.8	111	0.00
14 T	Allyl chloride	0.906	0.834	7.9	90	0.00
15 T	Acrylonitrile	0.345	0.366	-6.1	98	0.00
16 T	Acetone	0.306	0.297	2.9	95	0.00
17 T	Carbon Disulfide	1.509	1.308	13.3	92	0.00
18 T	Methyl Acetate	0.682	0.770	-12.9	106	0.00
19 T	Methyl tert-butyl Ether	1.912	1.902	0.5	90	0.00
20 T	Methylene Chloride	0.689	0.639	7.3	90	0.00
21 T	trans-1,2-Dichloroethene	0.620	0.569	8.2	92	0.00
22 T	Diisopropyl ether	1.850	1.845	0.3	91	0.00
23 T	Vinyl Acetate	1.577	1.581	-0.3	89	0.00
24 P	1,1-Dichloroethane	1.090	1.071	1.7	95	0.00
25 T	2-Butanone	0.445	0.467	-4.9	98	0.01
26 T	2,2-Dichloropropane	0.899	0.728	19.0	81	0.00
27 T	cis-1,2-Dichloroethene	0.699	0.667	4.6	91	0.00
28 T	Bromochloromethane	0.485	0.467	3.7	92	0.00
29 T	Tetrahydrofuran	0.280	0.300	-7.1	98	0.00
30 C	Chloroform	1.109	1.107	0.2#	94	0.00
31 T	Cyclohexane	0.924	0.873	5.5	101	0.00
32 T	1,1,1-Trichloroethane	0.925	0.941	-1.7	103	0.00
33 S	1,2-Dichloroethane-d4	0.670	0.658	1.8	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.320	0.323	-0.9	96	0.00
36 T	1,1-Dichloropropene	0.494	0.466	5.7	98	0.00
37 T	Ethyl Acetate	0.514	0.536	-4.3	96	0.00
38 T	Carbon Tetrachloride	0.470	0.480	-2.1	105	0.00
39 T	Methylcyclohexane	0.570	0.543	4.7	103	0.00
40 TM	Benzene	1.538	1.493	2.9	93	0.00
41 T	Methacrylonitrile	0.274	0.289	-5.5	93	0.00
42 TM	1,2-Dichloroethane	0.528	0.524	0.8	91	0.00
43 T	Isopropyl Acetate	0.811	0.833	-2.7	93	0.00
44 TM	Trichloroethene	0.398	0.376	5.5	95	0.00
45 C	1,2-Dichloropropane	0.389	0.389	0.0	93	0.00
46 T	Dibromomethane	0.266	0.260	2.3	91	0.00
47 T	Bromodichloromethane	0.507	0.532	-4.9	95	0.00
48 T	Methyl methacrylate	0.397	0.420	-5.8	94	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123120\
 Data File : VX020493.D
 Acq On : 31 Dec 2020 19:07
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 01 00:14:17 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50 S	Toluene-d8	1.231	1.215	1.3	95	0.00
51 T	4-Methyl-2-Pentanone	0.502	0.551	-9.8	98	0.00
52 CM	Toluene	0.939	0.932	0.7#	94	0.00
53 T	t-1,3-Dichloropropene	0.558	0.554	0.7	88	0.00
54 T	cis-1,3-Dichloropropene	0.612	0.605	1.1	88	0.00
55 T	1,1,2-Trichloroethane	0.387	0.396	-2.3	94	0.00
56 T	Ethyl methacrylate	0.567	0.600	-5.8	92	0.00
57 T	1,3-Dichloropropane	0.659	0.666	-1.1	92	0.00
58 T	2-Chloroethyl vinyl ether	0.300	0.286	4.7	86	0.00
59 T	2-Hexanone	0.382	0.426	-11.5	100	0.00
60 T	Dibromochloromethane	0.384	0.416	-8.3	97	0.00
61 T	1,2-Dibromoethane	0.404	0.414	-2.5	93	0.00
62 S	4-Bromofluorobenzene	0.471	0.486	-3.2	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.377	0.373	1.1	103	0.00
65 PM	Chlorobenzene	1.073	1.051	2.1	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.397	-3.1	96	0.00
67 C	Ethyl Benzene	1.875	1.879	-0.2#	95	0.00
68 T	m/p-Xylenes	0.700	0.713	-1.9	95	0.00
69 T	o-Xylene	0.674	0.684	-1.5	93	0.00
70 T	Styrene	1.139	1.202	-5.5	94	0.00
71 P	Bromoform	0.298	0.327	-9.7	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.609	3.532	2.1	102	0.00
74 T	N-aryl acetate	1.455	1.468	-0.9	96	0.00
75 P	1,1,2,2-Tetrachloroethane	1.310	1.327	-1.3	102	0.00
76 T	1,2,3-Trichloropropane	1.232	1.219	1.1	101	0.00
77 T	Bromobenzene	0.954	0.878	8.0	97	0.00
78 T	n-propylbenzene	4.159	4.067	2.2	102	0.00
79 T	2-Chlorotoluene	2.593	2.444	5.7	97	0.00
80 T	1,3,5-Trimethylbenzene	3.059	2.984	2.5	99	0.00
81 T	trans-1,4-Dichloro-2-butene	0.415	0.397	4.3	92	0.00
82 T	4-Chlorotoluene	3.029	2.855	5.7	96	0.00
83 T	tert-Butylbenzene	2.935	2.950	-0.5	105	0.00
84 T	1,2,4-Trimethylbenzene	3.105	3.072	1.1	98	0.00
85 T	sec-Butylbenzene	3.387	3.419	-0.9	109	0.00
86 T	p-Isopropyltoluene	3.124	3.097	0.9	103	0.00
87 T	1,3-Dichlorobenzene	1.705	1.605	5.9	96	0.00
88 T	1,4-Dichlorobenzene	1.779	1.607	9.7	95	0.00
89 T	n-Butylbenzene	2.737	2.717	0.7	106	0.00
90 T	Hexachloroethane	0.539	0.544	-0.9	108	0.00
91 T	1,2-Dichlorobenzene	1.697	1.617	4.7	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.275	0.287	-4.4	102	0.00
93 T	1,2,4-Trichlorobenzene	1.091	1.053	3.5	100	0.00
94 T	Hexachlorobutadiene	0.484	0.473	2.3	125	0.00
95 T	Naphthalene	3.456	3.607	-4.4	100	0.00
96 T	1,2,3-Trichlorobenzene	1.080	1.057	2.1	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123120\
Data File : VX020493.D
Acq On : 31 Dec 2020 19:07
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 01 00:14:17 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 5