

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123120\
 Data File : VX020493.D
 Acq On : 31 Dec 2020 19:07
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 01 00:14:17 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	99	0.01
2 T	Dichlorodifluoromethane	50.000	45.248	9.5	96	0.00
3 P	Chloromethane	50.000	44.182	11.6	89	0.00
4 C	Vinyl Chloride	50.000	45.177	9.6#	96	0.00
5 T	Bromomethane	50.000	49.336	1.3	103	0.00
6 T	Chloroethane	50.000	47.003	6.0	92	0.00
7 T	Trichlorofluoromethane	50.000	48.689	2.6	105	0.00
8 T	Diethyl Ether	50.000	48.154	3.7	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.932	4.1	104	0.00
10 T	Methyl Iodide	50.000	49.701	0.6	100	0.00
11 T	Tert butyl alcohol	250.000	252.458	-1.0	103	0.00
12 CM	1,1-Dichloroethene	50.000	47.036	5.9#	100	0.00
13 T	Acrolein	250.000	283.154	-13.3	111	0.00
14 T	Allyl chloride	50.000	46.018	8.0	90	0.00
15 T	Acrylonitrile	250.000	264.664	-5.9	98	0.00
16 T	Acetone	250.000	242.733	2.9	95	0.00
17 T	Carbon Disulfide	50.000	43.330	13.3	92	0.00
18 T	Methyl Acetate	50.000	56.504	-13.0	106	0.00
19 T	Methyl tert-butyl Ether	50.000	49.751	0.5	90	0.00
20 T	Methylene Chloride	50.000	46.379	7.2	90	0.00
21 T	trans-1,2-Dichloroethene	50.000	45.924	8.2	92	0.00
22 T	Diisopropyl ether	50.000	49.873	0.3	91	0.00
23 T	Vinyl Acetate	250.000	250.657	-0.3	89	0.00
24 P	1,1-Dichloroethane	50.000	49.139	1.7	95	0.00
25 T	2-Butanone	250.000	262.499	-5.0	98	0.01
26 T	2,2-Dichloropropane	50.000	40.502	19.0	81	0.00
27 T	cis-1,2-Dichloroethene	50.000	47.706	4.6	91	0.00
28 T	Bromochloromethane	50.000	48.136	3.7	92	0.00
29 T	Tetrahydrofuran	250.000	267.944	-7.2	98	0.00
30 C	Chloroform	50.000	49.872	0.3#	94	0.00
31 T	Cyclohexane	50.000	47.214	5.6	101	0.00
32 T	1,1,1-Trichloroethane	50.000	50.860	-1.7	103	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.105	1.8	93	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	95	0.00
35 S	Dibromofluoromethane	50.000	50.569	-1.1	96	0.00
36 T	1,1-Dichloropropene	50.000	47.111	5.8	98	0.00
37 T	Ethyl Acetate	50.000	52.151	-4.3	96	0.00
38 T	Carbon Tetrachloride	50.000	51.079	-2.2	105	0.00
39 T	Methylcyclohexane	50.000	47.568	4.9	103	0.00
40 TM	Benzene	50.000	48.534	2.9	93	0.00
41 T	Methacrylonitrile	50.000	52.829	-5.7	93	0.00
42 TM	1,2-Dichloroethane	50.000	49.638	0.7	91	0.00
43 T	Isopropyl Acetate	50.000	51.353	-2.7	93	0.00
44 TM	Trichloroethene	50.000	47.301	5.4	95	0.00
45 C	1,2-Dichloropropane	50.000	50.051	-0.1#	93	0.00
46 T	Dibromomethane	50.000	48.966	2.1	91	0.00
47 T	Bromodichloromethane	50.000	52.504	-5.0	95	0.00
48 T	Methyl methacrylate	50.000	52.796	-5.6	94	0.00
49 T	1,4-Dioxane	1000.000	1005.469	-0.5	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123120\
 Data File : VX020493.D
 Acq On : 31 Dec 2020 19:07
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 01 00:14:17 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	49.357	1.3	95	0.00
51 T	4-Methyl-2-Pentanone	250.000	274.544	-9.8	98	0.00
52 CM	Toluene	50.000	49.620	0.8#	94	0.00
53 T	t-1,3-Dichloropropene	50.000	49.629	0.7	88	0.00
54 T	cis-1,3-Dichloropropene	50.000	49.407	1.2	88	0.00
55 T	1,1,2-Trichloroethane	50.000	51.129	-2.3	94	0.00
56 T	Ethyl methacrylate	50.000	52.969	-5.9	92	0.00
57 T	1,3-Dichloropropane	50.000	50.551	-1.1	92	0.00
58 T	2-Chloroethyl vinyl ether	250.000	238.206	4.7	86	0.00
59 T	2-Hexanone	250.000	278.223	-11.3	100	0.00
60 T	Dibromochloromethane	50.000	54.129	-8.3	97	0.00
61 T	1,2-Dibromoethane	50.000	51.239	-2.5	93	0.00
62 S	4-Bromofluorobenzene	50.000	51.566	-3.1	99	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	92	0.00
64 T	Tetrachloroethene	50.000	49.570	0.9	103	0.00
65 PM	Chlorobenzene	50.000	48.991	2.0	93	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.521	-3.0	96	0.00
67 C	Ethyl Benzene	50.000	50.105	-0.2#	95	0.00
68 T	m/p-Xylenes	100.000	101.889	-1.9	95	0.00
69 T	o-Xylene	50.000	50.732	-1.5	93	0.00
70 T	Styrene	50.000	52.777	-5.6	94	0.00
71 P	Bromoform	50.000	54.785	-9.6	101	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	100	0.00
73 T	Isopropylbenzene	50.000	48.937	2.1	102	0.00
74 T	N-aryl acetate	50.000	50.471	-0.9	96	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.643	-1.3	102	0.00
76 T	1,2,3-Trichloropropane	50.000	49.503	1.0	101	0.00
77 T	Bromobenzene	50.000	46.022	8.0	97	0.00
78 T	n-propylbenzene	50.000	48.899	2.2	102	0.00
79 T	2-Chlorotoluene	50.000	47.131	5.7	97	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.783	2.4	99	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.757	4.5	92	0.00
82 T	4-Chlorotoluene	50.000	47.136	5.7	96	0.00
83 T	tert-Butylbenzene	50.000	50.245	-0.5	105	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.471	1.1	98	0.00
85 T	sec-Butylbenzene	50.000	50.461	-0.9	109	0.00
86 T	p-Isopropyltoluene	50.000	49.567	0.9	103	0.00
87 T	1,3-Dichlorobenzene	50.000	47.049	5.9	96	0.00
88 T	1,4-Dichlorobenzene	50.000	45.157	9.7	95	0.00
89 T	n-Butylbenzene	50.000	49.639	0.7	106	0.00
90 T	Hexachloroethane	50.000	50.506	-1.0	108	0.00
91 T	1,2-Dichlorobenzene	50.000	47.649	4.7	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	52.268	-4.5	102	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.246	3.5	100	0.00
94 T	Hexachlorobutadiene	50.000	51.487	-3.0	125	0.00
95 T	Naphthalene	50.000	52.192	-4.4	100	0.00
96 T	1,2,3-Trichlorobenzene	50.000	48.944	2.1	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123120\
Data File : VX020493.D
Acq On : 31 Dec 2020 19:07
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 01 00:14:17 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6