

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
 Data File : VX026157.D
 Acq On : 30 Dec 2021 10:42
 Operator : JC/MD
 Sample : VX1231MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1231MBL01

Quant Time: Dec 31 01:08:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	163002	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	274378	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236720	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99893	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	116048	49.211	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.420%
35) Dibromofluoromethane	5.391	113	91905	51.521	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.040%
50) Toluene-d8	8.653	98	319385	48.135	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.280%
62) 4-Bromofluorobenzene	11.079	95	107225	45.395	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
Data File : VX026157.D
Acq On : 30 Dec 2021 10:42
Operator : JC/MD
Sample : VX1231MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1231MBL01

Quant Time: Dec 31 01:08:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

