

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
 Data File : VX026172.D
 Acq On : 30 Dec 2021 16:46
 Operator : JC/MD
 Sample : M5207-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1S-122121

Quant Time: Dec 31 07:09:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	130547	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	224435	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198748	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87953	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99136	52.491	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 104.980%		
35) Dibromofluoromethane	5.391	113	76226	52.241	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.480%		
50) Toluene-d8	8.653	98	263506	48.551	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 97.100%		
62) 4-Bromofluorobenzene	11.085	95	91518	47.368	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 94.740%		
Target Compounds						Qvalue
17) Carbon Disulfide	2.514	76	1216	0.293	ug/l	97
27) cis-1,2-Dichloroethene	4.507	96	2225	1.240	ug/l	92
44) Trichloroethene	7.135	130	642	0.355	ug/l	77
64) Tetrachloroethene	9.281	164	13056	7.784	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
Data File : VX026172.D
Acq On : 30 Dec 2021 16:46
Operator : JC/MD
Sample : M5207-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1S-122121

Quant Time: Dec 31 07:09:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

