

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
 Data File : VX026173.D
 Acq On : 30 Dec 2021 17:10
 Operator : JC/MD
 Sample : M5207-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1D-122121

Quant Time: Dec 31 07:10:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	135792	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	228962	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202330	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84434	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99759	50.781	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.560%	
35) Dibromofluoromethane	5.391	113	77070	51.775	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.560%	
50) Toluene-d8	8.652	98	269068	48.596	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.200%	
62) 4-Bromofluorobenzene	11.085	95	91267	46.304	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.600%	
Target Compounds						
					Qvalue	
16) Acetone	2.391	43	2614	2.456	ug/l	97
17) Carbon Disulfide	2.513	76	1200	0.278	ug/l #	75
27) cis-1,2-Dichloroethene	4.501	96	6544	3.507	ug/l	97
44) Trichloroethene	7.128	130	1686	0.914	ug/l	83
64) Tetrachloroethene	9.274	164	18709	10.957	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
Data File : VX026173.D
Acq On : 30 Dec 2021 17:10
Operator : JC/MD
Sample : M5207-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1D-122121

Quant Time: Dec 31 07:10:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

