

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
 Data File : VX026175.D
 Acq On : 30 Dec 2021 17:56
 Operator : JC/MD
 Sample : M5207-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-1-122121

Quant Time: Dec 31 07:10:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132333	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	226320	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	199818	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85378	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99613	52.032	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.060%	
35) Dibromofluoromethane	5.391	113	76524	52.008	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.020%	
50) Toluene-d8	8.653	98	268810	49.116	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.240%	
62) 4-Bromofluorobenzene	11.085	95	91069	46.743	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.480%	
Target Compounds						
					Qvalue	
17) Carbon Disulfide	2.513	76	1277	0.304	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	2205	1.212	ug/l	96
44) Trichloroethene	7.135	130	823	0.452	ug/l	67
64) Tetrachloroethene	9.274	164	13173	7.812	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
Data File : VX026175.D
Acq On : 30 Dec 2021 17:56
Operator : JC/MD
Sample : M5207-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-1-122121

Quant Time: Dec 31 07:10:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

