

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
 Data File : VX026176.D
 Acq On : 30 Dec 2021 18:20
 Operator : JC/MD
 Sample : M5207-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-121321

Quant Time: Dec 31 07:10:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	129306	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	222821	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	195997	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85009	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	98196	52.492	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.980%	
35) Dibromofluoromethane	5.397	113	75287	51.971	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.940%	
50) Toluene-d8	8.653	98	263744	48.947	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.900%	
62) 4-Bromofluorobenzene	11.085	95	91899	47.909	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.820%	
Target Compounds						Qvalue
16) Acetone	2.386	43	2736	2.699	ug/l #	84
17) Carbon Disulfide	2.520	76	1350	0.329	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
Data File : VX026176.D
Acq On : 30 Dec 2021 18:20
Operator : JC/MD
Sample : M5207-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-121321

Quant Time: Dec 31 07:10:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

