

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
 Data File : VX026185.D
 Acq On : 30 Dec 2021 22:36
 Operator : JC/MD
 Sample : VX1231WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1231WBL02

Quant Time: Dec 31 07:12:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	137522	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	238803	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	211233	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92949	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	105507	53.031	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.060%
35) Dibromofluoromethane	5.397	113	81211	52.309	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.620%
50) Toluene-d8	8.653	98	283260	49.051	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.100%
62) 4-Bromofluorobenzene	11.085	95	97736	47.542	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
Data File : VX026185.D
Acq On : 30 Dec 2021 22:36
Operator : JC/MD
Sample : VX1231WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1231WBL02

Quant Time: Dec 31 07:12:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

