

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
 Data File : VX026191.D
 Acq On : 31 Dec 2021 00:55
 Operator : JC/MD
 Sample : PB141640ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141640ZHE#03

Quant Time: Dec 31 21:33:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	141305	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	243627	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	215575	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88281	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107628	52.649	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.300%
35) Dibromofluoromethane	5.397	113	82812	52.284	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.560%
50) Toluene-d8	8.653	98	289935	49.212	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.420%
62) 4-Bromofluorobenzene	11.085	95	97924	46.690	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.380%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	15370	13.877	ug/l	90
20) Methylene Chloride	2.794	84	8811	4.108	ug/l	93
43) Isopropyl Acetate	6.349	43	15690	3.250	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
Data File : VX026191.D
Acq On : 31 Dec 2021 00:55
Operator : JC/MD
Sample : PB141640ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141640ZHE#03

Quant Time: Dec 31 21:33:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

