

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
 Data File : VX026192.D
 Acq On : 31 Dec 2021 01:18
 Operator : JC/MD
 Sample : PB141694TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141694TB

Quant Time: Dec 31 21:33:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	139789	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	240701	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	222698	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97527	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	106889	52.854	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.700%	
35) Dibromofluoromethane	5.397	113	82466	52.698	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.400%	
50) Toluene-d8	8.653	98	293904	50.493	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.980%	
62) 4-Bromofluorobenzene	11.085	95	101379	48.925	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.860%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	16652	15.197	ug/l	97
20) Methylene Chloride	2.794	84	15591	8.107	ug/l	92
43) Isopropyl Acetate	6.348	43	15755	3.303	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
Data File : VX026192.D
Acq On : 31 Dec 2021 01:18
Operator : JC/MD
Sample : PB141694TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141694TB

Quant Time: Dec 31 21:33:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

