

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
 Data File : VX026199.D
 Acq On : 31 Dec 2021 04:01
 Operator : JC/MD
 Sample : PB141758TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141758TB

Quant Time: Dec 31 21:35:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	136032	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	239998	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	224100	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97167	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	107132	54.437	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 108.880%			
35) Dibromofluoromethane	5.391	113	82889	53.124	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.240%			
50) Toluene-d8	8.653	98	291456	50.219	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.440%			
62) 4-Bromofluorobenzene	11.079	95	103292	49.995	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 99.980%			
Target Compounds						
16) Acetone	2.386	43	9607	9.010	ug/l	89
20) Methylene Chloride	2.800	84	5092	2.081	ug/l	97
43) Isopropyl Acetate	6.349	43	16004	3.365	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
Data File : VX026199.D
Acq On : 31 Dec 2021 04:01
Operator : JC/MD
Sample : PB141758TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141758TB

Quant Time: Dec 31 21:35:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

