

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
 Data File : VX026200.D
 Acq On : 31 Dec 2021 04:25
 Operator : JC/MD
 Sample : PB141758ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141758ZHE#01

Quant Time: Dec 31 21:35:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	147134	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	254068	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231206	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96527	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	113905	53.512	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery	= 107.020%		
35) Dibromofluoromethane	5.391	113	88437	53.540	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 107.080%		
50) Toluene-d8	8.653	98	312277	50.826	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery	= 101.660%		
62) 4-Bromofluorobenzene	11.085	95	106325	48.613	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 97.220%		
Target Compounds						Qvalue
16) Acetone	2.386	43	11872	10.294	ug/l	93
20) Methylene Chloride	2.794	84	9638	4.364	ug/l #	86
43) Isopropyl Acetate	6.348	43	15372	3.053	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
Data File : VX026200.D
Acq On : 31 Dec 2021 04:25
Operator : JC/MD
Sample : PB141758ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141758ZHE#01

Quant Time: Dec 31 21:35:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

