

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
 Data File : VX026201.D
 Acq On : 31 Dec 2021 04:48
 Operator : JC/MD
 Sample : PB141758ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141758ZHE#02

Quant Time: Dec 31 21:36:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	138717	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	239575	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219587	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88988	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107938	53.785	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 107.580%			
35) Dibromofluoromethane	5.397	113	82944	53.253	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.500%			
50) Toluene-d8	8.653	98	295005	50.920	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.840%			
62) 4-Bromofluorobenzene	11.085	95	96377	46.730	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 93.460%			
Target Compounds						
16) Acetone	2.386	43	11363	10.450	ug/l	94
20) Methylene Chloride	2.794	84	9445	4.574	ug/l #	85
43) Isopropyl Acetate	6.348	43	15144	3.190	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
Data File : VX026201.D
Acq On : 31 Dec 2021 04:48
Operator : JC/MD
Sample : PB141758ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141758ZHE#02

Quant Time: Dec 31 21:36:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

